

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV041520\
 Data File : VV015509.D
 Acq On : 15 Apr 2020 15:12
 Operator : SY/MD
 Sample : L2273-04
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 YAQ82

Quant Time: Apr 16 01:37:31 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR040920WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Apr 16 00:42:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	113824	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	105620	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	51728	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	29808	5.04	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	100.80%
7) Chloroethane-d5	1.58	69	27393	5.04	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	100.80%
11) 1,1-Dichloroethene-d2	2.12	63	43260	3.48	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.60%
20) 2-Butanone-d5	3.92	46	77386	46.67	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	93.34%
24) Chloroform-d	4.38	84	58765	4.49	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.80%
26) 1,2-Dichloroethane-d4	5.06	65	33715	4.78	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.60%
32) Benzene-d6	5.08	84	117758	5.03	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.60%
36) 1,2-Dichloropropane-d6	6.09	67	36497	5.00	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.00%
41) Toluene-d8	7.34	98	105937	4.74	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.80%
43) trans-1,3-Dichloropropene-	7.65	79	14743	4.64	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	92.80%
46) 2-Hexanone-d5	8.11	63	60535	39.55	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	79.10%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	24174	4.17	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	83.40%
64) 1,2-Dichlorobenzene-d4	11.65	152	38576	4.65	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	93.00%

Target Compounds					Ovalue
13) Acetone	2.20	43	4886	4.170 ug/L #	50
25) Chloroform	4.41	83	4259	0.274 ug/L	98
47) Tetrachloroethene	8.00	164	21650	3.434 ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV041520\
Data File : VV015509.D
Acq On : 15 Apr 2020 15:12
Operator : SY/MD
Sample : L2273-04
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
YAQ82

Quant Time: Apr 16 01:37:31 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR040920WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Apr 16 00:42:12 2020
Response via : Initial Calibration

