

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV041520\
 Data File : VV015516.D
 Acq On : 15 Apr 2020 18:18
 Operator : SY/MD
 Sample : L2274-15DL 10X
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 YAQB3DL

Quant Time: Apr 16 02:09:23 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR040920WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Apr 16 00:42:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	109301	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	100877	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	49060	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	28839	5.08	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	101.60%
7) Chloroethane-d5	1.58	69	25519	4.89	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	97.80%
11) 1,1-Dichloroethene-d2	2.12	63	40658	3.40	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	68.00%
20) 2-Butanone-d5	3.94	46	77330	48.56	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	97.12%
24) Chloroform-d	4.38	84	57269	4.56	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.20%
26) 1,2-Dichloroethane-d4	5.06	65	32606	4.82	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.40%
32) Benzene-d6	5.08	84	114542	5.12	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.40%
36) 1,2-Dichloropropane-d6	6.09	67	36049	5.17	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	103.40%
41) Toluene-d8	7.34	98	103076	4.83	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.60%
43) trans-1,3-Dichloropropene-	7.64	79	13830	4.56	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.20%
46) 2-Hexanone-d5	8.11	63	62020	42.42	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	84.84%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	23929	4.33	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	86.60%
64) 1,2-Dichlorobenzene-d4	11.65	152	37453	4.76	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	95.20%

Target Compounds				Ovalue
47) Tetrachloroethene	8.00	164	44800	7.441 ug/L 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV041520\
Data File : VV015516.D
Acq On : 15 Apr 2020 18:18
Operator : SY/MD
Sample : L2274-15DL 10X
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
YAQB3DL

Quant Time: Apr 16 02:09:23 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR040920WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Apr 16 00:42:12 2020
Response via : Initial Calibration

