

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV041520\
 Data File : VV015508.D
 Acq On : 15 Apr 2020 14:49
 Operator : SY/MD
 Sample : L2273-02DL 10X
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 YAQ80DL

Quant Time: Apr 16 01:34:07 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR040920WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Apr 16 00:42:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	110520	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	102794	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	49793	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	28990	5.05	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	101.00%
7) Chloroethane-d5	1.58	69	26539	5.03	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	100.60%
11) 1,1-Dichloroethene-d2	2.12	63	41593	3.44	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	68.80%
20) 2-Butanone-d5	3.93	46	77142	47.91	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	95.82%
24) Chloroform-d	4.38	84	57667	4.54	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.80%
26) 1,2-Dichloroethane-d4	5.06	65	32935	4.81	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.20%
32) Benzene-d6	5.08	84	115163	5.05	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.00%
36) 1,2-Dichloropropane-d6	6.10	67	36291	5.11	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	102.20%
41) Toluene-d8	7.34	98	104685	4.81	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.20%
43) trans-1,3-Dichloropropene-	7.65	79	14182	4.59	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.80%
46) 2-Hexanone-d5	8.11	63	61999	41.62	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	83.24%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	24459	4.34	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	86.80%
64) 1,2-Dichlorobenzene-d4	11.65	152	38393	4.80	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.00%


Target Compounds					Ovalue
47) Tetrachloroethene	8.00	164	56336	9.182 ug/L	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV041520\
Data File : VV015508.D
Acq On : 15 Apr 2020 14:49
Operator : SY/MD
Sample : L2273-02DL 10X
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
YAQ80DL

Quant Time: Apr 16 01:34:07 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR040920WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Apr 16 00:42:12 2020
Response via : Initial Calibration





#47

Tetrachloroethene

Concen: 9.182 ug/L

RT: 8.00 min Scan# 2148

Delta R.T. 0.00 min

Lab File: VV015508.D


Acq: 15 Apr 2020 14:49

Instrument :

MSVOA_V

ClientSampleId :

YAQ80DL



Tot Ion:164 Resp: 56336

Ion Ratio Lower Upper

164 100

129 98.6 70.6 131.0

131 95.0 68.3 126.8

166 128.1 88.3 163.9

Abundance Ion 163.90 (163.60 to 164.60): V

Ion 128.95 (128.65 to 129.65): V

Ion 130.95 (130.65 to 131.65): V

Ion 165.90 (165.60 to 166.60): V

