

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV041620\
 Data File : VV015535.D
 Acq On : 16 Apr 2020 14:38
 Operator : SY/MD
 Sample : L2320-02
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFRQ9

Quant Time: Apr 17 01:48:14 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR040920WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Apr 17 00:46:36 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	105962	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	99995	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	49294	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	26815	4.88	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	97.60%
7) Chloroethane-d5	1.57	69	24434	4.83	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	96.60%
11) 1,1-Dichloroethene-d2	2.12	63	38363	3.31	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	66.20%
20) 2-Butanone-d5	3.96	46	75806	49.11	ug/L	0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	98.22%
24) Chloroform-d	4.38	84	54960	4.51	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.20%
26) 1,2-Dichloroethane-d4	5.06	65	32456	4.95	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.00%
32) Benzene-d6	5.07	84	110696	4.99	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.80%
36) 1,2-Dichloropropane-d6	6.10	67	34965	5.06	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.20%
41) Toluene-d8	7.34	98	99362	4.69	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.80%
43) trans-1,3-Dichloropropene-	7.65	79	14001	4.66	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	93.20%
46) 2-Hexanone-d5	8.12	63	62501	43.13	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	86.26%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	22503	4.10	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	82.00%
64) 1,2-Dichlorobenzene-d4	11.65	152	36984	4.67	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	93.40%

Target Compounds

					Ovalue
17) Methyl tert-butyl Ether	2.79	73	5449	0.307 ug/L	98
34) Trichloroethene	5.95	95	1593	0.198 ug/L	94
47) Tetrachloroethene	8.00	164	887	0.149 ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV041620\
Data File : VV015535.D
Acq On : 16 Apr 2020 14:38
Operator : SY/MD
Sample : L2320-02
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFRQ9

Quant Time: Apr 17 01:48:14 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR040920WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Apr 17 00:46:36 2020
Response via : Initial Calibration

