

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW041621\
 Data File : VW021046.D
 Acq On : 16 Apr 2021 14:39
 Operator : SY/MD
 Sample : M2067-04
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG799

Manual Integrations
 APPROVED

MMDadoda
 4/20/2021 10:33:00 AM

Quant Time: Apr 17 00:49:20 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM041421WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Apr 17 00:47:19 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.622	114	418730	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.857	117	412941	50.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.255	152	208832	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.310	65	97860	42.96	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	85.92%
7) Chloroethane-d5	1.571	69	89125	49.77	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	99.54%
11) 1,1-Dichloroethene-d2	2.111	63	174456	35.70	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.40%
21) 2-Butanone-d5	3.908	46	164274	109.65	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	109.65%
24) Chloroform-d	4.352	84	297966	52.69	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.38%
26) 1,2-Dichloroethane-d4	5.037	65	210509	55.08	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	110.16%
32) Benzene-d6	5.056	84	537567	52.73	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.46%
36) 1,2-Dichloropropane-d6	6.076	67	152342	53.82	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	107.64%
41) Toluene-d8	7.320	98	521642	51.63	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.26%
43) trans-1,3-Dichloroprop...	7.625	79	91608	52.52	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	105.04%
47) 2-Hexanone-d5	8.095	63	146938	120.86	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	120.86%
56) 1,1,2,2-Tetrachloroeth...	10.220	84	238040	51.89	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	103.78%
66) 1,2-Dichlorobenzene-d4	11.632	152	247967	58.16	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	116.32%
Target Compounds						
46) Tetrachloroethene	7.989	164	16993722m	6009.54	ug/L	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041621\
Data File : VV021046.D
Acq On : 16 Apr 2021 14:39
Operator : SY/MD
Sample : M2067-04
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1


Instrument :
MSVOA_V
ClientSampleId :
BG799

Manual Integrations
APPROVED

MMDadoda
4/20/2021 10:33:00 AM


Quant Time: Apr 17 00:49:20 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM041421WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Apr 17 00:47:19 2021
Response via : Initial Calibration





#46
Tetrachloroethene
Concen: 6009.54 ug/L m
RT: 7.989 min Scan# 2161
Delta R.T. 0.010 min
Lab File: VV021046.D
Acq: 16 Apr 2021 14:39

Instrument :
MSVOA_V
ClientSampleId :
BG799



Tgt Ion:164 Resp:16993722
Ion Ratio Lower Upper
164 100
129 97.5 66.5 123.5
131 95.2 63.8 118.4
166 100.0 93.4 173.4

Manual Integrations
APPROVED

MMDadoda
4/20/2021 10:33:00 AM

