

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW041624\
 Data File : VW035233.D
 Acq On : 16 Apr 2024 12:25
 Operator : SY/MD
 Sample : P2100-01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COAN8

Quant Time: Apr 17 02:48:08 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR041524WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Apr 17 02:28:28 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	191310	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	197334	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	82902	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	80018	4.467	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	89.400%
7) Chloroethane-d5	1.532	69	67516	4.851	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	97.000%
11) 1,1-Dichloroethene-d2	2.060	65	34508	4.541	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	90.800%
20) 2-Butanone-d5	3.815	46	169364	55.897	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	111.800%
24) Chloroform-d	4.249	84	136763	4.883	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.600%
26) 1,2-Dichloroethane-d4	4.944	65	69388	5.205	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.000%
32) Benzene-d6	4.960	84	265408	4.734	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.600%
36) 1,2-Dichloropropane-d6	5.989	67	89303	5.122	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	102.400%
41) Toluene-d8	7.243	98	214791	4.266	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.400%
43) trans-1,3-Dichloroprop...	7.558	79	26820	4.310	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	86.200%
46) 2-Hexanone-d5	8.027	63	116677	45.022	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	90.040%
56) 1,1,2,2-Tetrachloroeth...	10.153	84	64906	4.833	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.600%
66) 1,2-Dichlorobenzene-d4	11.561	152	77913	5.287	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.800%
Target Compounds						
12) 1,1-Dichloroethene	2.069	96	9459	0.651	ug/L	# 1
19) 1,1-Dichloroethane	3.114	63	13413	0.449	ug/L	98
22) cis-1,2-Dichloroethene	3.815	96	141778	9.292	ug/L	99
25) Chloroform	4.285	83	6004	0.207	ug/L	92
34) Trichloroethene	5.828	95	1014412	62.170	ug/L	99
47) Tetrachloroethene	7.905	164	28920	2.234	ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041624\
Data File : VV035233.D
Acq On : 16 Apr 2024 12:25
Operator : SY/MD
Sample : P2100-01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
COAN8

Quant Time: Apr 17 02:48:08 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR041524WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Apr 17 02:28:28 2024
Response via : Initial Calibration

