

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW041624\
 Data File : VW035242.D
 Acq On : 16 Apr 2024 15:57
 Operator : SY/MD
 Sample : P2117-01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E0MH4

Quant Time: Apr 17 02:50:11 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR041524WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Apr 17 02:28:28 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	214498	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	218364	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	94482	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	78449	3.906	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	78.200%
7) Chloroethane-d5	1.532	69	68229	4.372	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	87.400%
11) 1,1-Dichloroethene-d2	2.060	65	33736	3.959	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	79.200%
20) 2-Butanone-d5	3.818	46	142206	41.860	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	83.720%
24) Chloroform-d	4.249	84	138869	4.422	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.400%
26) 1,2-Dichloroethane-d4	4.944	65	69647	4.660	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.200%
32) Benzene-d6	4.960	84	261787	4.220	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.400%
36) 1,2-Dichloropropane-d6	5.989	67	87385	4.530	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	90.600%
41) Toluene-d8	7.243	98	212422	3.812	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	76.200%
43) trans-1,3-Dichloroprop...	7.561	79	25402	3.689	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	73.800%
46) 2-Hexanone-d5	8.027	63	111718	38.957	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	77.920%
56) 1,1,2,2-Tetrachloroeth...	10.153	84	65432	4.403	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	88.000%
66) 1,2-Dichlorobenzene-d4	11.561	152	78276	4.661	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	93.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041624\
Data File : VV035242.D
Acq On : 16 Apr 2024 15:57
Operator : SY/MD
Sample : P2117-01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E0MH4

Quant Time: Apr 17 02:50:11 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR041524WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Apr 17 02:28:28 2024
Response via : Initial Calibration

