

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW041624\
 Data File : VW035247.D
 Acq On : 16 Apr 2024 17:57
 Operator : SY/MD
 Sample : P2117-07
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E0MH3

Quant Time: Apr 17 02:51:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR041524WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Apr 17 02:28:28 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	182368	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	189225	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	83004	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	71738	4.201	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	84.000%
7) Chloroethane-d5	1.532	69	61396	4.628	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	92.600%
11) 1,1-Dichloroethene-d2	2.060	65	31640	4.368	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	87.400%
20) 2-Butanone-d5	3.815	46	148161	51.297	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	102.600%
24) Chloroform-d	4.249	84	126502	4.738	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.800%
26) 1,2-Dichloroethane-d4	4.944	65	65060	5.120	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.400%
32) Benzene-d6	4.960	84	244672	4.551	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.000%
36) 1,2-Dichloropropane-d6	5.989	67	80115	4.792	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.800%
41) Toluene-d8	7.243	98	198419	4.110	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.200%
43) trans-1,3-Dichloroprop...	7.561	79	23189	3.887	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	77.800%
46) 2-Hexanone-d5	8.027	63	104642	42.109	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	84.220%
56) 1,1,2,2-Tetrachloroeth...	10.153	84	61625	4.786	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.800%
66) 1,2-Dichlorobenzene-d4	11.561	152	72690	4.927	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.600%
Target Compounds						
22) cis-1,2-Dichloroethene	3.818	96	9831	0.676	ug/L	88
27) 1,2-Dichloroethane	5.063	62	6106	0.376	ug/L #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041624\
Data File : VV035247.D
Acq On : 16 Apr 2024 17:57
Operator : SY/MD
Sample : P2117-07
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E0MH3

Quant Time: Apr 17 02:51:15 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR041524WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Apr 17 02:28:28 2024
Response via : Initial Calibration

