

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW041624\
 Data File : VW035241.D
 Acq On : 16 Apr 2024 15:33
 Operator : SY/MD
 Sample : P2032-22
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E0MG6

Quant Time: Apr 17 02:49:59 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR041524WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Apr 17 02:28:28 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	183872	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	189638	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	82391	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	79258	4.603	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	92.000%
7) Chloroethane-d5	1.532	69	66527	4.973	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	99.400%
11) 1,1-Dichloroethene-d2	2.060	65	34287	4.694	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	93.800%
20) 2-Butanone-d5	3.812	46	164053	56.335	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	112.660%
24) Chloroform-d	4.249	84	137224	5.097	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.000%
26) 1,2-Dichloroethane-d4	4.944	65	70898	5.533	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	110.600%
32) Benzene-d6	4.960	84	263109	4.884	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.600%
36) 1,2-Dichloropropane-d6	5.989	67	87511	5.223	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.400%
41) Toluene-d8	7.243	98	213037	4.403	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.000%
43) trans-1,3-Dichloroprop...	7.558	79	27094	4.531	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.600%
46) 2-Hexanone-d5	8.027	63	119408	47.946	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	95.900%
56) 1,1,2,2-Tetrachloroeth...	10.152	84	66248	5.133	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	102.600%
66) 1,2-Dichlorobenzene-d4	11.561	152	79218	5.409	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	108.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041624\
Data File : VV035241.D
Acq On : 16 Apr 2024 15:33
Operator : SY/MD
Sample : P2032-22
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E0MG6

Quant Time: Apr 17 02:49:59 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR041524WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Apr 17 02:28:28 2024
Response via : Initial Calibration

