

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV041720\
 Data File : VV015555.D
 Acq On : 17 Apr 2020 15:14
 Operator : SY/MD
 Sample : L2288-02
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFCK1

Quant Time: Apr 18 01:54:02 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR040920WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Apr 18 00:37:52 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	0.00	114	0	0.00	ug/L	-5.64
28) Chlorobenzene-d5	8.88	117	401	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.28	152	570	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	0.00	65	0	0.00	ug/L	
Spiked Amount	5.000	Range	40 - 130	Recovery	=	0.00%#
7) Chloroethane-d5	1.85	69	20	0.00	ug/L	0.27
Spiked Amount	5.000	Range	65 - 130	Recovery	=	0.00%#
11) 1,1-Dichloroethene-d2	2.12	63	174	0.00	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	0.00%#
20) 2-Butanone-d5	0.00	46	0	0.00	ug/L	
Spiked Amount	50.000	Range	40 - 130	Recovery	=	0.00%#
24) Chloroform-d	4.40	84	156	0.00	ug/L	0.03
Spiked Amount	5.000	Range	70 - 125	Recovery	=	0.00%#
26) 1,2-Dichloroethane-d4	0.00	65	0	0.00	ug/L	
Spiked Amount	5.000	Range	70 - 130	Recovery	=	0.00%#
32) Benzene-d6	0.00	84	0	0.00	ug/L	
Spiked Amount	5.000	Range	70 - 125	Recovery	=	0.00%#
36) 1,2-Dichloropropane-d6	5.94	67	19	0.69	ug/L	-0.15
Spiked Amount	5.000	Range	60 - 140	Recovery	=	13.80%#
41) Toluene-d8	0.00	98	0	0.00	ug/L	
Spiked Amount	5.000	Range	70 - 130	Recovery	=	0.00%#
43) trans-1,3-Dichloropropene-	0.00	79	0	0.00	ug/L	
Spiked Amount	5.000	Range	55 - 130	Recovery	=	0.00%#
46) 2-Hexanone-d5	0.00	63	0	0.00	ug/L	
Spiked Amount	50.000	Range	45 - 130	Recovery	=	0.00%#
57) 1,1,2,2-Tetrachloroethane-	0.00	84	0	0.00	ug/L	
Spiked Amount	5.000	Range	65 - 120	Recovery	=	0.00%#
64) 1,2-Dichlorobenzene-d4	11.65	152	330	3.61	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	72.20%#

Target Compounds

					Ovalue
33) Benzene	5.13	78	694	5.782	ug/L 100
34) Trichloroethene	5.94	95	29494	913.014	ug/L 99
35) Methylcyclohexane	5.96	83	136	2.499	ug/L # 15
40) 4-Methyl-2-pentanone	7.17	43	20	1.023	ug/L # 46
42) Toluene	7.42	91	289	2.218	ug/L 89
47) Tetrachloroethene	8.00	164	9142	381.971	ug/L 95
48) 2-Hexanone	8.20	43	174	12.741	ug/L # 40
49) Dibromochloromethane	8.00	129	8835	368.600	ug/L # 10
62) 1,3-Dichlorobenzene	11.30	146	157	0.854	ug/L # 26
63) 1,4-Dichlorobenzene	11.30	146	157	0.840	ug/L # 25
69) Naphthalene	13.53	128	20	0.089	ug/L # 69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV041720\
Data File : VV015555.D
Acq On : 17 Apr 2020 15:14
Operator : SY/MD
Sample : L2288-02
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFCK1

Quant Time: Apr 18 01:54:02 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR040920WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Apr 18 00:37:52 2020
Response via : Initial Calibration

