

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV041720\
 Data File : VV015568.D
 Acq On : 17 Apr 2020 20:21
 Operator : SY/MD
 Sample : L2333-11
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AC3

Quant Time: Apr 18 03:56:40 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR040920WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Apr 18 00:37:52 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	92955	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	88932	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	44003	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	21327	4.42	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	88.40%
7) Chloroethane-d5	1.58	69	21874	4.93	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	98.60%
11) 1,1-Dichloroethene-d2	2.12	63	33080	3.26	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.20%
20) 2-Butanone-d5	3.91	46	79511	58.71	ug/L	-0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	117.42%
24) Chloroform-d	4.38	84	50416	4.72	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.40%
26) 1,2-Dichloroethane-d4	5.06	65	30041	5.22	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.40%
32) Benzene-d6	5.08	84	95108	4.82	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.40%
36) 1,2-Dichloropropane-d6	6.09	67	31242	5.08	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.60%
41) Toluene-d8	7.34	98	85015	4.51	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.20%
43) trans-1,3-Dichloropropene-	7.64	79	8618	3.22	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	64.40%
46) 2-Hexanone-d5	8.11	63	62900	48.80	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	97.60%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	22785	4.67	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.40%
64) 1,2-Dichlorobenzene-d4	11.65	152	32989	4.67	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	93.40%

Target Compounds

					Ovalue
13) Acetone	2.18	43	7499	7.838 ug/L #	48
21) 2-Butanone	3.99	43	44760	28.949 ug/L #	67
22) cis-1,2-Dichloroethene	3.94	96	1798	0.263 ug/L	84
34) Trichloroethene	5.94	95	3545	0.495 ug/L	95
42) Toluene	7.41	91	6098	0.211 ug/L	98
47) Tetrachloroethene	8.00	164	597	0.112 ug/L	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV041720\
Data File : VV015568.D
Acq On : 17 Apr 2020 20:21
Operator : SY/MD
Sample : L2333-11
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AC3

Quant Time: Apr 18 03:56:40 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR040920WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Apr 18 00:37:52 2020
Response via : Initial Calibration

