

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW041724\
 Data File : VW035259.D
 Acq On : 17 Apr 2024 11:29
 Operator : SY/MD
 Sample : P2100-02DL 10X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AN9DL

Quant Time: Apr 18 00:41:43 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR041524WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Apr 18 00:40:19 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	190714	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	194772	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	82276	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	70032	3.922	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	78.400%
7) Chloroethane-d5	1.532	69	60082	4.330	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	86.600%
11) 1,1-Dichloroethene-d2	2.056	65	30442	4.018	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	80.400%
20) 2-Butanone-d5	3.828	46	152996	50.653	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	101.300%
24) Chloroform-d	4.249	84	127882	4.580	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.600%
26) 1,2-Dichloroethane-d4	4.944	65	65326	4.916	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.400%
32) Benzene-d6	4.960	84	243023	4.392	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.800%
36) 1,2-Dichloropropane-d6	5.989	67	81185	4.718	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.400%
41) Toluene-d8	7.243	98	192867	3.881	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	77.600%
43) trans-1,3-Dichloroprop...	7.561	79	25310	4.121	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	82.400%
46) 2-Hexanone-d5	8.030	63	104665	40.918	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	81.840%
56) 1,1,2,2-Tetrachloroeth...	10.153	84	61500	4.640	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.800%
66) 1,2-Dichlorobenzene-d4	11.558	152	71524	4.891	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	97.800%
Target Compounds						
					Qvalue	
16) Methylene chloride	2.449	84	16528	0.906	ug/L	97
22) cis-1,2-Dichloroethene	3.825	96	12242	0.805	ug/L	97
34) Trichloroethene	5.831	95	82758	5.139	ug/L	98
47) Tetrachloroethene	7.912	164	2590	0.203	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041724\
Data File : VV035259.D
Acq On : 17 Apr 2024 11:29
Operator : SY/MD
Sample : P2100-02DL 10X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AN9DL

Quant Time: Apr 18 00:41:43 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR041524WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Apr 18 00:40:19 2024
Response via : Initial Calibration

