

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW041724\
 Data File : VW035262.D
 Acq On : 17 Apr 2024 12:39
 Operator : SY/MD
 Sample : P2100-07
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Apr 18 00:42:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR041524WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Apr 18 00:40:19 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	188723	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	200054	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	86312	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	76500	4.329	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	86.600%
7) Chloroethane-d5	1.532	69	66387	4.835	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	96.800%
11) 1,1-Dichloroethene-d2	2.060	65	31688	4.227	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	84.600%
20) 2-Butanone-d5	3.828	46	164842	55.151	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	110.300%
24) Chloroform-d	4.249	84	136730	4.949	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.000%
26) 1,2-Dichloroethane-d4	4.947	65	70192	5.337	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.800%
32) Benzene-d6	4.963	84	262034	4.611	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.200%
36) 1,2-Dichloropropane-d6	5.992	67	87131	4.930	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.600%
41) Toluene-d8	7.246	98	210228	4.118	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.400%
43) trans-1,3-Dichloroprop...	7.561	79	25120	3.982	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	79.600%
46) 2-Hexanone-d5	8.031	63	115668	44.026	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	88.060%
56) 1,1,2,2-Tetrachloroeth...	10.153	84	66423	4.879	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.600%
66) 1,2-Dichlorobenzene-d4	11.561	152	81336	5.301	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	106.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041724\
Data File : VV035262.D
Acq On : 17 Apr 2024 12:39
Operator : SY/MD
Sample : P2100-07
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Apr 18 00:42:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR041524WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Apr 18 00:40:19 2024
Response via : Initial Calibration

