

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV041819\
 Data File : VV010400.D
 Acq On : 19 Apr 2019 23:10
 Operator : SY/MD
 Sample : K2456-17
 Misc : 5.10G/10mL/MSVOA V/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GAHN7

Quant Time: Apr 20 05:26:18 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM041819S.M
 Quant Title : VOC Analysis
 QLast Update : Sat Apr 20 04:30:57 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	216494	25.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	211484	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	107863	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	54087	19.67	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	78.68%
7) Chloroethane-d5	1.58	69	50780	21.99	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	87.96%
10) 1,1-Dichloroethene-d2	2.13	63	87633	13.21	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	52.84%
20) 2-Butanone-d5	3.94	46	55779	65.15	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	130.30%
24) Chloroform-d	4.40	84	124078	21.80	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	87.20%
26) 1,2-Dichloroethane-d4	5.08	65	89559	26.66	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	106.64%
29) Benzene-d6	5.10	84	221511	20.46	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	81.84%
33) 1,2-Dichloropropane-d6	6.12	67	71932	22.70	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	90.80%
37) Toluene-d8	7.36	98	214270	20.59	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	82.36%
38) trans-1,3-Dichloropropene-	7.66	79	33280	21.36	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	85.44%
39) 2-Hexanone-d5	8.13	63	41672	59.07	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	118.14%
48) 1,1,2,2-Tetrachloroethane-	10.26	84	85372	26.94	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	107.76%
61) 1,2-Dichlorobenzene-d4	11.67	152	102813	24.24	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	96.96%

Target Compounds					Ovalue
13) Acetone	2.19	43	3608	3.993 ug/L	99
16) Methylene chloride	2.54	84	14913	4.313 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV041819\
Data File : VV010400.D
Acq On : 19 Apr 2019 23:10
Operator : SY/MD
Sample : K2456-17
Misc : 5.10G/10mL/MSVOA V/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GAHN7

Quant Time: Apr 20 05:26:18 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM041819S.M
Quant Title : VOC Analysis
QLast Update : Sat Apr 20 04:30:57 2019
Response via : Initial Calibration

