

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW041823\
 Data File : VW030644.D
 Acq On : 19 Apr 2023 02:15
 Operator : SY/MD
 Sample : 02333-16 200X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COMH1

Quant Time: Apr 19 04:31:05 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM041723WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Apr 18 02:45:56 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	230642	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.792	117	224543	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	104360	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	54107	32.001	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	64.000%
7) Chloroethane-d5	1.889	69	166	0.117	ug/L	0.35
Spiked Amount	50.000	Range	70 - 130	Recovery	=	0.240%#
11) 1,1-Dichloroethene-d2	2.066	65	28891	36.032	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	72.060%
21) 2-Butanone-d5	3.799	46	79829	85.960	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	85.960%
24) Chloroform-d	4.262	84	119179	38.895	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	77.800%
26) 1,2-Dichloroethane-d4	4.950	65	82853	42.058	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	84.120%
32) Benzene-d6	4.973	84	209343	39.433	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	78.860%
36) 1,2-Dichloropropane-d6	5.995	67	69651	42.237	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	84.480%
41) Toluene-d8	7.252	98	181198	36.352	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	72.700%#
43) trans-1,3-Dichloroprop...	7.561	79	29796	35.849	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.700%
47) 2-Hexanone-d5	8.034	63	51728	79.627	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	79.630%
56) 1,1,2,2-Tetrachloroeth...	10.162	84	99306	41.501	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	83.000%
66) 1,2-Dichlorobenzene-d4	11.570	152	76656	46.213	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.420%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041823\
Data File : VV030644.D
Acq On : 19 Apr 2023 02:15
Operator : SY/MD
Sample : 02333-16 200X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
COMH1

Quant Time: Apr 19 04:31:05 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM041723WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Apr 18 02:45:56 2023
Response via : Initial Calibration

