

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW041823\
 Data File : VW030637.D
 Acq On : 18 Apr 2023 23:28
 Operator : SY/MD
 Sample : VW0418WBL02
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK324

Quant Time: Apr 19 04:29:56 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM041723WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Apr 18 02:45:56 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	231828	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.792	117	226699	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	105315	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	69433	40.855	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	81.720%
7) Chloroethane-d5	1.535	69	58359	40.978	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	81.960%
11) 1,1-Dichloroethene-d2	2.066	65	34868	43.263	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	86.520%
21) 2-Butanone-d5	3.796	46	82564	88.450	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	88.450%
24) Chloroform-d	4.259	84	137573	44.669	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.340%
26) 1,2-Dichloroethane-d4	4.950	65	93157	47.047	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.100%
32) Benzene-d6	4.969	84	246318	45.956	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.920%
36) 1,2-Dichloropropane-d6	5.998	67	78468	47.131	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	94.260%
41) Toluene-d8	7.252	98	216214	42.965	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	85.920%
43) trans-1,3-Dichloroprop...	7.561	79	36546	43.552	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.100%
47) 2-Hexanone-d5	8.034	63	52841	80.567	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	80.570%
56) 1,1,2,2-Tetrachloroeth...	10.162	84	105869	43.823	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	87.640%
66) 1,2-Dichlorobenzene-d4	11.570	152	84517	50.490	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.980%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041823\
Data File : VV030637.D
Acq On : 18 Apr 2023 23:28
Operator : SY/MD
Sample : VV0418WBL02
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK324

Quant Time: Apr 19 04:29:56 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM041723WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Apr 18 02:45:56 2023
Response via : Initial Calibration

