

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041923\
 Data File : VV030673.D
 Acq On : 19 Apr 2023 18:30
 Operator : SY/MD
 Sample : 02334-02
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COMH5

Quant Time: Apr 20 01:30:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM041723WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Apr 20 01:26:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	220568	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.793	117	214826	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	100883	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.282	65	53702	33.212	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	66.420%
7) Chloroethane-d5	1.536	69	52442	38.703	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	77.400%
11) 1,1-Dichloroethene-d2	2.066	65	27580	35.968	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.940%
21) 2-Butanone-d5	3.802	46	90622	102.038	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	102.040%
24) Chloroform-d	4.259	84	131870	45.003	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.000%
26) 1,2-Dichloroethane-d4	4.950	65	93809	49.795	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.580%
32) Benzene-d6	4.970	84	230068	45.297	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.600%
36) 1,2-Dichloropropane-d6	5.998	67	76507	48.493	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.980%
41) Toluene-d8	7.252	98	194190	40.721	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	81.440%
43) trans-1,3-Dichloroprop...	7.561	79	35679	44.869	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	89.740%
47) 2-Hexanone-d5	8.034	63	59847	96.292	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	96.290%
56) 1,1,2,2-Tetrachloroeth...	10.159	84	114785	50.139	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	100.280%
66) 1,2-Dichlorobenzene-d4	11.567	152	87822	54.769	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	109.540%
Target Compounds						
					Qvalue	
51) Chlorobenzene	8.821	112	7040	1.521	ug/L	89
65) 1,4-Dichlorobenzene	11.220	146	4880	1.469	ug/L	92
67) 1,2-Dichlorobenzene	11.587	146	3488	1.098	ug/L #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041923\
Data File : VV030673.D
Acq On : 19 Apr 2023 18:30
Operator : SY/MD
Sample : 02334-02
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
COMH5

Quant Time: Apr 20 01:30:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM041723WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Apr 20 01:26:42 2023
Response via : Initial Calibration

