

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV042020\
 Data File : VV015583.D
 Acq On : 20 Apr 2020 16:04
 Operator : SY/MD
 Sample : L2333-05
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AB7

Quant Time: Apr 21 02:53:40 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR040920WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Apr 21 00:05:52 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	99066	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	92697	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	45722	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	21793	4.24	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	84.80%
7) Chloroethane-d5	1.58	69	21867	4.62	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	92.40%
11) 1,1-Dichloroethene-d2	2.12	63	34086	3.15	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	63.00%
20) 2-Butanone-d5	3.95	46	80025	55.45	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	110.90%
24) Chloroform-d	4.38	84	53309	4.68	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.60%
26) 1,2-Dichloroethane-d4	5.06	65	31456	5.13	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.60%
32) Benzene-d6	5.07	84	100519	4.89	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.80%
36) 1,2-Dichloropropane-d6	6.09	67	32567	5.09	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.80%
41) Toluene-d8	7.34	98	89529	4.56	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.20%
43) trans-1,3-Dichloropropene-	7.65	79	10980	3.94	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	78.80%
46) 2-Hexanone-d5	8.11	63	64016	47.65	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	95.30%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	23292	4.58	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	91.60%
64) 1,2-Dichlorobenzene-d4	11.65	152	34421	4.69	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	93.80%

Target Compounds

					Ovalue
25) Chloroform	4.41	83	2185	0.162 ug/L	94
31) Carbon tetrachloride	4.85	117	2666	0.255 ug/L	99
34) Trichloroethene	5.94	95	3963	0.531 ug/L	95
42) Toluene	7.41	91	2714	0.090 ug/L	89
47) Tetrachloroethene	8.00	164	3450	0.624 ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV042020\
Data File : VV015583.D
Acq On : 20 Apr 2020 16:04
Operator : SY/MD
Sample : L2333-05
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AB7

Quant Time: Apr 21 02:53:40 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR040920WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Apr 21 00:05:52 2020
Response via : Initial Calibration

