

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV042020\
 Data File : VV015588.D
 Acq On : 20 Apr 2020 18:01
 Operator : SY/MD
 Sample : L2333-09
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AC1

Manual Integrations
 APPROVED

apatel
 4/21/2020 10:44:07 AM

Quant Time: Apr 21 03:40:40 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR040920WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Apr 21 00:05:52 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	102305	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	96873	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	47425	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	21525	4.05	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	81.00%
7) Chloroethane-d5	1.58	69	21952	4.49	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.80%
11) 1,1-Dichloroethene-d2	2.12	63	34702	3.10	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	62.00%
20) 2-Butanone-d5	3.92	46	81635	54.77	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	109.54%
24) Chloroform-d	4.38	84	58008	4.93	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.60%
26) 1,2-Dichloroethane-d4	5.06	65	31016	4.90	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.00%
32) Benzene-d6	5.08	84	100151	4.66	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.20%
36) 1,2-Dichloropropane-d6	6.10	67	33304	4.98	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.60%
41) Toluene-d8	7.34	98	91395	4.46	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.20%
43) trans-1,3-Dichloropropene-	7.64	79	10809	3.71	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	74.20%
46) 2-Hexanone-d5	8.11	63	62886	44.79	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	89.58%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	22529	4.24	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	84.80%
64) 1,2-Dichlorobenzene-d4	11.65	152	35222	4.63	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	92.60%

Target Compounds

					Ovalue
13) Acetone	2.20	43	4152m	3.943	ug/L
14) Carbon disulfide	2.31	76	9014	0.403	ug/L
25) Chloroform	4.41	83	96997	6.948	ug/L
38) Bromodichloromethane	6.54	83	5125	0.534	ug/L
42) Toluene	7.41	91	4022	0.128	ug/L

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV042020\
Data File : VV015588.D
Acq On : 20 Apr 2020 18:01
Operator : SY/MD
Sample : L2333-09
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AC1

Manual Integrations
APPROVED

apatel
4/21/2020 10:44:07 AM

Quant Time: Apr 21 03:40:40 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR040920WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Apr 21 00:05:52 2020
Response via : Initial Calibration

