


Data Path : Z:\voasrv\HPCHEM1\MSVOA V\Data\VV042020\
 Data File : VV015570.D
 Acq On : 20 Apr 2020 09:30
 Operator : SY/MD
 Sample : BFB38
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFB38

Integration File: rteint.p

Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR040920WMA.M
 Title : TRACE VOA SOM01.0
 Last Update : Tue Apr 21 00:05:52 2020



AutoFind: Scans 2800, 2801, 2802; Background Corrected with Scan 2788

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.1	14965	PASS
75	95	30	80	54.4	38616	PASS
95	95	100	100	100.0	71018	PASS
96	95	5	9	6.9	4920	PASS
173	174	0.00	2	0.9	461	PASS
174	95	50	120	72.7	51648	PASS
175	174	5	9	7.5	3890	PASS
176	174	95	101	95.1	49138	PASS
177	176	5	9	6.7	3281	PASS

m/z	Abundance
36.00	700.0
37.00	3757.0
38.00	3573.0
39.00	1480.0
39.90	306.0
42.90	104.0
44.00	736.0
45.00	700.0
47.00	839.0
47.80	542.0
49.00	3373.0
50.00	14633.0
51.00	4415.0
52.00	280.0
55.00	228.0
56.00	1054.0
57.00	2173.0
58.10	114.0
59.90	630.0
61.00	3816.0
62.00	3786.0
63.00	2819.0
64.00	326.0
65.00	383.0
66.90	260.0
68.00	7458.0
69.00	7904.0
70.00	677.0
72.00	414.0
73.00	3402.0
74.00	12554.0
75.00	37448.0
76.00	3297.0
76.90	391.0
77.90	223.0
78.90	2539.0
80.00	745.0
80.90	2607.0
81.80	536.0
85.80	127.0
86.90	2663.0
87.90	2594.0
90.80	339.0
92.00	2077.0
93.00	3235.0
94.00	8335.0
95.00	68632.0
96.00	4747.0
97.00	190.0
103.90	292.0
104.80	132.0
105.80	297.0
110.80	111.0
115.80	251.0
116.90	514.0
117.80	317.0
118.80	366.0
127.90	306.0
128.90	116.0
129.80	276.0
130.90	148.0
134.90	123.0
136.70	164.0
140.90	641.0
142.80	665.0
144.80	168.0
145.80	136.0
147.80	166.0
154.90	149.0
158.90	112.0
171.90	338.0
173.00	668.0
173.90	48224.0
175.00	3695.0
176.00	45800.0
176.90	3362.0

Instrument :
MSVOA_V
ClientSampleId :
BFB38

m/z	Abundance
36.00	737.0
37.00	3988.0
38.00	3789.0
39.00	1566.0
40.00	319.0
44.00	796.0
45.00	699.0
47.00	1022.0
47.90	560.0
49.00	3619.0
50.00	15415.0
51.00	4682.0
52.00	301.0
55.00	243.0
56.00	1173.0
57.00	2268.0
58.20	100.0
59.90	715.0
61.00	3987.0
62.00	4020.0
63.00	3047.0
64.00	342.0
65.00	364.0
66.90	272.0
68.00	7882.0
69.00	8426.0
70.00	678.0
72.00	433.0
73.00	3636.0
74.00	13552.0
75.00	39816.0
76.00	3496.0
76.90	387.0
77.80	265.0
78.90	2709.0
80.00	805.0
80.90	2910.0
81.80	598.0
85.90	125.0
86.90	2769.0
87.90	2780.0
90.80	328.0
92.00	2206.0
93.00	3402.0
94.00	8986.0
95.00	73288.0
96.00	5099.0
97.00	198.0
103.80	331.0
104.80	128.0
105.80	324.0
112.70	102.0
115.90	275.0
116.90	529.0
117.80	318.0
118.80	419.0
127.90	298.0
128.90	126.0
129.80	319.0
130.90	139.0
134.80	149.0
136.70	165.0
140.90	697.0
142.90	754.0
144.80	144.0
145.80	157.0
147.80	179.0
154.80	150.0
156.90	101.0
158.90	119.0
171.90	353.0
173.00	716.0
173.90	53392.0
175.00	4048.0
176.00	50624.0
176.90	3405.0

Instrument :
MSVOA_V
ClientSampleId :
BFB38

m/z	Abundance
36.00	722.0
37.00	3888.0
38.00	3685.0
39.00	1425.0
40.00	318.0
44.00	804.0
45.00	660.0
47.00	1012.0
47.90	527.0
49.00	3467.0
50.00	14847.0
51.00	4497.0
51.90	254.0
55.00	258.0
56.00	1194.0
57.00	2098.0
59.90	726.0
61.00	3860.0
62.00	3945.0
63.00	2956.0
64.00	314.0
65.00	309.0
66.90	251.0
68.00	7660.0
69.00	8033.0
70.00	584.0
71.90	463.0
73.00	3573.0
74.00	13316.0
75.00	38584.0
76.00	3410.0
77.00	370.0
77.80	281.0
78.90	2673.0
80.00	827.0
80.90	2904.0
81.80	606.0
82.90	104.0
86.90	2685.0
87.90	2749.0
90.90	277.0
92.00	2176.0
93.00	3275.0
94.00	8791.0
95.00	71136.0
96.00	4916.0
97.00	163.0
103.80	349.0
104.90	102.0
105.80	351.0
112.80	105.0
115.90	298.0
116.90	524.0
117.80	275.0
118.80	422.0
127.90	273.0
128.80	120.0
129.80	322.0
130.80	104.0
134.70	167.0
136.80	146.0
140.90	691.0
142.90	825.0
144.80	112.0
145.80	166.0
147.90	175.0
154.80	145.0
158.90	112.0
172.00	368.0
174.00	53328.0
175.00	3927.0
175.90	50992.0
176.90	3076.0

Instrument :
MSVOA_V
ClientSampleId :
BFB38

Scan 2788 (10.055 min): VV015570.D
BFB38

m/z	Abundance
39.90	312.0
44.00	287.0

Instrument :
MSVOA_V
ClientSampleId :
BFB38