

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW042023\
 Data File : VW030733.D
 Acq On : 20 Apr 2023 22:04
 Operator : SY/MD
 Sample : 02411-20
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Apr 21 06:12:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM041723WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Apr 21 01:37:39 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	238454	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.792	117	228284	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	112013	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	71271	40.772	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	81.540%
7) Chloroethane-d5	1.535	69	58454	39.904	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	79.800%
11) 1,1-Dichloroethene-d2	2.063	65	33561	40.485	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	80.960%
21) 2-Butanone-d5	3.796	46	85070	88.602	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	88.600%
24) Chloroform-d	4.259	84	129782	40.968	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	81.940%
26) 1,2-Dichloroethane-d4	4.950	65	93606	45.960	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.920%
32) Benzene-d6	4.969	84	238541	44.197	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.400%
36) 1,2-Dichloropropane-d6	5.998	67	72750	43.393	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	86.780%
41) Toluene-d8	7.252	98	211374	41.712	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	83.420%
43) trans-1,3-Dichloroprop...	7.561	79	35018	41.441	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	82.880%
47) 2-Hexanone-d5	8.034	63	54116	81.938	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	81.940%
56) 1,1,2,2-Tetrachloroeth...	10.159	84	102145	41.988	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	83.980%
66) 1,2-Dichlorobenzene-d4	11.570	152	86413	48.536	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.080%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042023\
Data File : VV030733.D
Acq On : 20 Apr 2023 22:04
Operator : SY/MD
Sample : 02411-20
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Apr 21 06:12:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM041723WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Apr 21 01:37:39 2023
Response via : Initial Calibration

