

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW042023\
 Data File : VW030737.D
 Acq On : 20 Apr 2023 23:40
 Operator : SY/MD
 Sample : 02440-06 100X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COMQ1

Quant Time: Apr 21 06:13:44 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM041723WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Apr 21 01:37:39 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	226629	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.792	117	216651	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	105492	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	67929	40.887	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	81.780%
7) Chloroethane-d5	1.535	69	56111	40.303	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	80.600%
11) 1,1-Dichloroethene-d2	2.063	65	33962	43.106	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	86.220%
21) 2-Butanone-d5	3.796	46	85447	93.638	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	93.640%
24) Chloroform-d	4.259	84	128498	42.679	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.360%
26) 1,2-Dichloroethane-d4	4.950	65	91794	47.422	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.840%
32) Benzene-d6	4.969	84	234956	45.870	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.740%
36) 1,2-Dichloropropane-d6	5.998	67	72673	45.675	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	91.340%
41) Toluene-d8	7.252	98	205698	42.771	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	85.540%
43) trans-1,3-Dichloroprop...	7.561	79	35000	43.644	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.280%
47) 2-Hexanone-d5	8.034	63	54214	86.494	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	86.490%
56) 1,1,2,2-Tetrachloroeth...	10.162	84	101138	43.806	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	87.620%
66) 1,2-Dichlorobenzene-d4	11.570	152	85828	51.187	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.380%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042023\
Data File : VV030737.D
Acq On : 20 Apr 2023 23:40
Operator : SY/MD
Sample : 02440-06 100X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
COMQ1

Quant Time: Apr 21 06:13:44 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM041723WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Apr 21 01:37:39 2023
Response via : Initial Calibration

