

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW042023\
 Data File : VW030725.D
 Acq On : 20 Apr 2023 18:33
 Operator : SY/MD
 Sample : 02411-18
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AS1

Quant Time: Apr 21 00:20:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM041723WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Apr 21 00:14:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	240276	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.792	117	233847	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	108989	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	72366	41.084	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	82.160%
7) Chloroethane-d5	1.539	69	61476	41.649	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	83.300%
11) 1,1-Dichloroethene-d2	2.066	65	33405	39.991	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	79.980%
21) 2-Butanone-d5	3.809	46	81197	83.927	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	83.930%
24) Chloroform-d	4.262	84	129930	40.704	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	81.400%
26) 1,2-Dichloroethane-d4	4.953	65	90543	44.119	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.240%
32) Benzene-d6	4.969	84	242242	43.815	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.620%
36) 1,2-Dichloropropane-d6	5.998	67	73636	42.877	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	85.760%
41) Toluene-d8	7.252	98	211040	40.655	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	81.300%
43) trans-1,3-Dichloroprop...	7.561	79	35326	40.811	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	81.620%
47) 2-Hexanone-d5	8.034	63	53611	79.242	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	79.240%
56) 1,1,2,2-Tetrachloroeth...	10.159	84	101192	40.606	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	81.220%
66) 1,2-Dichlorobenzene-d4	11.570	152	82838	47.819	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.640%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042023\
Data File : VV030725.D
Acq On : 20 Apr 2023 18:33
Operator : SY/MD
Sample : 02411-18
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AS1

Quant Time: Apr 21 00:20:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM041723WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Apr 21 00:14:50 2023
Response via : Initial Calibration

