

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV042120\
 Data File : VV015609.D
 Acq On : 21 Apr 2020 18:37
 Operator : SY/MD
 Sample : L2366-14
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 ETKJ0

Quant Time: Apr 22 03:21:45 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR040920WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Apr 22 00:11:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	135273	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	128909	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	63538	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	33883	4.83	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	96.60%
7) Chloroethane-d5	1.58	69	32599	5.04	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	100.80%
11) 1,1-Dichloroethene-d2	2.12	63	54072	3.66	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	73.20%
20) 2-Butanone-d5	3.92	46	106507	54.04	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	108.08%
24) Chloroform-d	4.38	84	74687	4.80	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.00%
26) 1,2-Dichloroethane-d4	5.06	65	44425	5.30	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.00%
32) Benzene-d6	5.08	84	146136	5.11	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.20%
36) 1,2-Dichloropropane-d6	6.09	67	46271	5.20	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.00%
41) Toluene-d8	7.34	98	134680	4.93	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.60%
43) trans-1,3-Dichloropropene-	7.64	79	14322	3.70	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	74.00%
46) 2-Hexanone-d5	8.11	63	83104	44.48	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	88.96%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	31993	4.53	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.60%
64) 1,2-Dichlorobenzene-d4	11.65	152	51106	5.01	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV042120\
Data File : VV015609.D
Acq On : 21 Apr 2020 18:37
Operator : SY/MD
Sample : L2366-14
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
ETKJ0

Quant Time: Apr 22 03:21:45 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR040920WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Apr 22 00:11:50 2020
Response via : Initial Calibration

