

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW042121\
 Data File : VW021077.D
 Acq On : 21 Apr 2021 14:49
 Operator : SY/MD
 Sample : VIBLK24
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK24

Quant Time: Apr 22 06:42:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM042121WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Apr 22 06:32:19 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.622	114	524926	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.857	117	507807	50.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.255	152	238122	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	254897	59.10	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	118.20%
7) Chloroethane-d5	1.571	69	191989	60.51	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	121.02%
11) 1,1-Dichloroethene-d2	2.108	63	358913	45.11	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	90.22%
21) 2-Butanone-d5	3.905	46	255079	121.10	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	121.10%
24) Chloroform-d	4.352	84	447790	58.23	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	116.46%
26) 1,2-Dichloroethane-d4	5.037	65	306837	62.77	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	125.54%#
32) Benzene-d6	5.053	84	842223	62.28	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	124.56%
36) 1,2-Dichloropropane-d6	6.072	67	252012	62.43	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	124.86%#
41) Toluene-d8	7.320	98	787013	60.13	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	120.26%#
43) trans-1,3-Dichloroprop...	7.625	79	129569	58.73	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	117.46%
47) 2-Hexanone-d5	8.095	63	206386	128.90	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	128.90%
56) 1,1,2,2-Tetrachloroeth...	10.223	84	354284	58.19	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	116.38%
66) 1,2-Dichlorobenzene-d4	11.632	152	322970	68.66	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	137.32%#
Target Compounds						
18) Methyl tert-butyl Ether	2.773	73	53221	4.82	ug/L	91
34) Trichloroethene	5.924	95	11427	2.92	ug/L	94
46) Tetrachloroethene	7.979	164	80592	23.00	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042121\
Data File : VV021077.D
Acq On : 21 Apr 2021 14:49
Operator : SY/MD
Sample : VIBLK24
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK24

Quant Time: Apr 22 06:42:11 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM042121WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Apr 22 06:32:19 2021
Response via : Initial Calibration

