

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042123\
 Data File : VV030756.D
 Acq On : 21 Apr 2023 11:45
 Operator : SY/MD
 Sample : VV0421WBL01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK330

Quant Time: Apr 22 00:44:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM041723WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Apr 22 00:43:23 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	242615	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.792	117	232581	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	116058	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	69958	39.334	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	78.660%
7) Chloroethane-d5	1.539	69	60378	40.511	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	81.020%
11) 1,1-Dichloroethene-d2	2.066	65	33799	40.072	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	80.140%
21) 2-Butanone-d5	3.805	46	84062	86.051	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	86.050%
24) Chloroform-d	4.259	84	133453	41.404	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	82.800%
26) 1,2-Dichloroethane-d4	4.953	65	95715	46.189	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.380%
32) Benzene-d6	4.969	84	245557	44.656	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.320%
36) 1,2-Dichloropropane-d6	5.995	67	76795	44.959	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	89.920%
41) Toluene-d8	7.249	98	219685	42.551	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	85.100%
43) trans-1,3-Dichloroprop...	7.558	79	37192	43.201	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	86.400%
47) 2-Hexanone-d5	8.034	63	52249	77.649	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	77.650%
56) 1,1,2,2-Tetrachloroeth...	10.159	84	104460	42.146	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	84.300%
66) 1,2-Dichlorobenzene-d4	11.567	152	91469	49.585	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.160%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042123\
Data File : VV030756.D
Acq On : 21 Apr 2023 11:45
Operator : SY/MD
Sample : VV0421WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK330

Quant Time: Apr 22 00:44:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM041723WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Apr 22 00:43:23 2023
Response via : Initial Calibration

