

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV042220\
 Data File : VV015628.D
 Acq On : 22 Apr 2020 18:02
 Operator : SY/MD
 Sample : L2366-22MSD
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 ETKJ9MSD

Quant Time: Apr 23 03:31:03 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Apr 23 03:05:46 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	138267	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	129831	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	64406	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	33804	4.68	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	93.60%
7) Chloroethane-d5	1.57	69	31689	4.55	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	91.00%
11) 1,1-Dichloroethene-d2	2.12	63	76239	4.82	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	96.40%
20) 2-Butanone-d5	3.95	46	102231	44.06	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	88.12%
24) Chloroform-d	4.38	84	74806	4.44	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.80%
26) 1,2-Dichloroethane-d4	5.06	65	43264	4.59	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.80%
32) Benzene-d6	5.07	84	145947	4.71	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.20%
36) 1,2-Dichloropropane-d6	6.09	67	45617	4.61	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.20%
41) Toluene-d8	7.34	98	134517	4.57	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.40%
43) trans-1,3-Dichloropropene-	7.64	79	13940	3.64	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	72.80%
46) 2-Hexanone-d5	8.11	63	82149	43.99	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	87.98%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	33136	4.49	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	89.80%
64) 1,2-Dichlorobenzene-d4	11.65	152	50953	4.65	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	93.00%

Target Compounds

					Ovalue
12) 1,1-Dichloroethene	2.13	96	41494	5.156	ug/L 95
33) Benzene	5.13	78	201042	5.420	ug/L 100
34) Trichloroethene	5.94	95	52414	5.117	ug/L 97
42) Toluene	7.41	91	210705	5.316	ug/L 100
51) Chlorobenzene	8.90	112	136163	5.244	ug/L 99
53) m,p-xylene	9.16	106	1280	0.076	ug/L 79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV042220\
Data File : VV015628.D
Acq On : 22 Apr 2020 18:02
Operator : SY/MD
Sample : L2366-22MSD
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
ETKJ9MSD

Quant Time: Apr 23 03:31:03 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Apr 23 03:05:46 2020
Response via : Initial Calibration

