

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV042220\
 Data File : VV015635.D
 Acq On : 22 Apr 2020 20:46
 Operator : SY/MD
 Sample : L2366-18
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 ETKJ4

Quant Time: Apr 23 05:02:19 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Apr 23 03:05:46 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	127700	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	124720	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	59425	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	31914	4.78	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	95.60%
7) Chloroethane-d5	1.58	69	32002	4.98	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	99.60%
11) 1,1-Dichloroethene-d2	2.12	63	50477	3.46	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.20%
20) 2-Butanone-d5	3.91	46	105949	49.44	ug/L	-0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	98.88%
24) Chloroform-d	4.38	84	74689	4.80	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.00%
26) 1,2-Dichloroethane-d4	5.06	65	43427	4.99	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.80%
32) Benzene-d6	5.08	84	144480	4.85	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.00%
36) 1,2-Dichloropropane-d6	6.09	67	45905	4.83	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.60%
41) Toluene-d8	7.34	98	132500	4.69	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.80%
43) trans-1,3-Dichloropropene-	7.64	79	14632	3.98	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	79.60%
46) 2-Hexanone-d5	8.11	63	79767	44.47	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	88.94%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	31731	4.48	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	89.60%
64) 1,2-Dichlorobenzene-d4	11.65	152	50599	5.01	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.20%

Target Compounds

					Ovalue
13) Acetone	2.19	43	14443	9.707	ug/L 99
16) Methylene chloride	2.53	84	1776	0.194	ug/L 96
52) Ethylbenzene	9.04	91	6299	0.143	ug/L 95
53) m,p-xylene	9.16	106	9101	0.564	ug/L 93
54) o-xylene	9.57	106	4106	0.263	ug/L 81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV042220\
Data File : VV015635.D
Acq On : 22 Apr 2020 20:46
Operator : SY/MD
Sample : L2366-18
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
ETKJ4

Quant Time: Apr 23 05:02:19 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Apr 23 03:05:46 2020
Response via : Initial Calibration

