

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV042220\
 Data File : VV015645.D
 Acq On : 23 Apr 2020 00:41
 Operator : SY/MD
 Sample : L2370-09
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 ETKK5

Quant Time: Apr 23 05:49:01 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Apr 23 03:05:46 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	122898	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	119393	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	56854	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	30970	4.82	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	96.40%
7) Chloroethane-d5	1.58	69	30088	4.86	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	97.20%
11) 1,1-Dichloroethene-d2	2.12	63	48628	3.46	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.20%
20) 2-Butanone-d5	3.91	46	95750	46.43	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	92.86%
24) Chloroform-d	4.38	84	70770	4.73	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.60%
26) 1,2-Dichloroethane-d4	5.06	65	41109	4.91	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.20%
32) Benzene-d6	5.08	84	134811	4.73	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.60%
36) 1,2-Dichloropropane-d6	6.09	67	42162	4.63	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.60%
41) Toluene-d8	7.34	98	122840	4.54	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.80%
43) trans-1,3-Dichloropropene-	7.65	79	11294	3.21	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	64.20%
46) 2-Hexanone-d5	8.11	63	73638	42.88	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	85.76%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	30448	4.49	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	89.80%
64) 1,2-Dichlorobenzene-d4	11.65	152	46468	4.81	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.20%

Target Compounds					Ovalue
17) Methyl tert-butyl Ether	2.78	73	8220	0.409 ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV042220\
Data File : VV015645.D
Acq On : 23 Apr 2020 00:41
Operator : SY/MD
Sample : L2370-09
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
ETKK5

Quant Time: Apr 23 05:49:01 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Apr 23 03:05:46 2020
Response via : Initial Calibration

