

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW042221\
 Data File : VW021107.D
 Acq On : 22 Apr 2021 20:22
 Operator : SY/MD
 Sample : M2065-11ME
 Misc : 2.65g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG564ME

Quant Time: Apr 23 04:36:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM042121WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Apr 23 04:33:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	722610	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.857	117	690275	50.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.255	152	342451	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	223098	37.57	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	75.14%		
7) Chloroethane-d5	1.568	69	175678	40.22	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	80.44%		
11) 1,1-Dichloroethene-d2	2.105	63	343753	31.39	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	62.78%		
21) 2-Butanone-d5	3.915	46	251956	86.90	ug/L	0.02
Spiked Amount 100.000	Range 40 - 130		Recovery =	86.90%		
24) Chloroform-d	4.352	84	440908	41.65	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	83.30%		
26) 1,2-Dichloroethane-d4	5.037	65	309332	45.97	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	91.94%		
32) Benzene-d6	5.053	84	826789	44.97	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	89.94%		
36) 1,2-Dichloropropane-d6	6.072	67	252752	46.06	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	92.12%		
41) Toluene-d8	7.320	98	763260	42.90	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	85.80%		
43) trans-1,3-Dichloroprop...	7.625	79	123869	41.31	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	82.62%		
47) 2-Hexanone-d5	8.095	63	206540	94.90	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	94.90%		
56) 1,1,2,2-Tetrachloroeth...	10.220	84	353242	42.68	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	85.36%		
66) 1,2-Dichlorobenzene-d4	11.631	152	318062	47.02	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	94.04%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042221\
Data File : VV021107.D
Acq On : 22 Apr 2021 20:22
Operator : SY/MD
Sample : M2065-11ME
Misc : 2.65g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BG564ME

Quant Time: Apr 23 04:36:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM042121WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Apr 23 04:33:55 2021
Response via : Initial Calibration

