

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW042221\
 Data File : VW021105.D
 Acq On : 22 Apr 2021 19:35
 Operator : SY/MD
 Sample : M2065-15ME
 Misc : 2.61g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG568ME

Quant Time: Apr 23 04:36:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM042121WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Apr 23 04:33:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	662013	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.857	117	650739	50.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.255	152	313732	50.00	ug/L	0.00

System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	221333	40.69	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	81.38%
7) Chloroethane-d5	1.565	69	169991	42.48	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	84.96%
11) 1,1-Dichloroethene-d2	2.101	63	330141	32.90	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	65.80%
21) 2-Butanone-d5	3.918	46	220567	83.03	ug/L	0.02
Spiked Amount	100.000	Range	40 - 130	Recovery	=	83.03%
24) Chloroform-d	4.349	84	435000	44.85	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.70%
26) 1,2-Dichloroethane-d4	5.034	65	300839	48.80	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.60%
32) Benzene-d6	5.050	84	815646	47.06	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.12%
36) 1,2-Dichloropropane-d6	6.072	67	248750	48.09	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.18%
41) Toluene-d8	7.320	98	753171	44.90	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	89.80%
43) trans-1,3-Dichloroprop...	7.625	79	126518	44.75	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	89.50%
47) 2-Hexanone-d5	8.095	63	205254	100.04	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	100.04%
56) 1,1,2,2-Tetrachloroeth...	10.220	84	353358	45.29	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	90.58%
66) 1,2-Dichlorobenzene-d4	11.632	152	312371	50.40	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.80%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042221\
Data File : VV021105.D
Acq On : 22 Apr 2021 19:35
Operator : SY/MD
Sample : M2065-15ME
Misc : 2.61g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BG568ME

Quant Time: Apr 23 04:36:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM042121WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Apr 23 04:33:55 2021
Response via : Initial Calibration

