

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042224\
 Data File : VV035324.D
 Acq On : 22 Apr 2024 11:49
 Operator : SY/MD
 Sample : VV0422WBL01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK272

Quant Time: Apr 23 00:28:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Apr 23 00:28:07 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	414869	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	409915	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	186655	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	203860	50.881	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	101.760%
7) Chloroethane-d5	1.535	69	164063	47.144	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	94.280%
11) 1,1-Dichloroethene-d2	2.059	65	83972	47.812	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	95.620%
21) 2-Butanone-d5	3.793	46	220709	92.882	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	92.880%
24) Chloroform-d	4.246	84	340110	50.488	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.980%
26) 1,2-Dichloroethane-d4	4.937	65	213938	50.022	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.040%
32) Benzene-d6	4.956	84	697203	50.635	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.260%
36) 1,2-Dichloropropane-d6	5.985	67	221457	50.649	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	101.300%
41) Toluene-d8	7.239	98	612759	49.415	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.820%
43) trans-1,3-Dichloroprop...	7.551	79	94796	44.597	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	89.200%
47) 2-Hexanone-d5	8.024	63	120012	72.597	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	72.600%
56) 1,1,2,2-Tetrachloroeth...	10.149	84	292439	48.569	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	97.140%
66) 1,2-Dichlorobenzene-d4	11.557	152	225677	53.850	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	107.700%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042224\
Data File : VV035324.D
Acq On : 22 Apr 2024 11:49
Operator : SY/MD
Sample : VV0422WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK272

Quant Time: Apr 23 00:28:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Apr 23 00:28:07 2024
Response via : Initial Calibration

