

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV042320\
 Data File : VV015662.D
 Acq On : 23 Apr 2020 15:02
 Operator : SY/MD
 Sample : L2393-03
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Apr 24 05:40:08 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Apr 24 03:28:15 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	130302	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	124348	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	60520	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	29505	4.33	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	86.60%
7) Chloroethane-d5	1.57	69	29332	4.47	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.40%
11) 1,1-Dichloroethene-d2	2.11	63	46935	3.15	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	63.00%
20) 2-Butanone-d5	3.95	46	104284	47.69	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	95.38%
24) Chloroform-d	4.38	84	70101	4.42	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.40%
26) 1,2-Dichloroethane-d4	5.06	65	43284	4.87	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.40%
32) Benzene-d6	5.07	84	138484	4.67	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.40%
36) 1,2-Dichloropropane-d6	6.09	67	43955	4.63	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.60%
41) Toluene-d8	7.34	98	124356	4.41	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.20%
43) trans-1,3-Dichloropropene-	7.64	79	14607	3.98	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	79.60%
46) 2-Hexanone-d5	8.11	63	81164	45.38	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	90.76%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	31505	4.46	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	89.20%
64) 1,2-Dichlorobenzene-d4	11.65	152	49701	4.83	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.60%

Target Compounds					Ovalue
3) Chloromethane	1.24	50	1526	0.158 ug/L	97
25) Chloroform	4.40	83	4709	0.266 ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV042320\
Data File : VV015662.D
Acq On : 23 Apr 2020 15:02
Operator : SY/MD
Sample : L2393-03
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Apr 24 05:40:08 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Apr 24 03:28:15 2020
Response via : Initial Calibration

