

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW042321\
 Data File : VW021129.D
 Acq On : 23 Apr 2021 19:23
 Operator : SY/MD
 Sample : M2123-01 100X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG7B2

Quant Time: Apr 24 03:47:20 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM042121WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Apr 24 03:42:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.622	114	689539	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.857	117	674671	50.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.255	152	310424	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	176506	31.15	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	62.30%
7) Chloroethane-d5	1.571	69	147618	35.42	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	70.84%
11) 1,1-Dichloroethene-d2	2.108	63	274241	26.24	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	52.48%#
21) 2-Butanone-d5	3.912	46	261470	94.50	ug/L	0.02
Spiked Amount	100.000	Range	40 - 130	Recovery	=	94.50%
24) Chloroform-d	4.352	84	395622	39.17	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	78.34%
26) 1,2-Dichloroethane-d4	5.037	65	285120	44.40	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.80%
32) Benzene-d6	5.053	84	719690	40.05	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	80.10%
36) 1,2-Dichloropropane-d6	6.072	67	228819	42.66	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	85.32%
41) Toluene-d8	7.320	98	647517	37.24	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	74.48%#
43) trans-1,3-Dichloroprop...	7.625	79	114816	39.17	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	78.34%
47) 2-Hexanone-d5	8.095	63	204922	96.33	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	96.33%
56) 1,1,2,2-Tetrachloroeth...	10.220	84	341390	42.20	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	84.40%
66) 1,2-Dichlorobenzene-d4	11.631	152	279399	45.56	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.12%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042321\
Data File : VV021129.D
Acq On : 23 Apr 2021 19:23
Operator : SY/MD
Sample : M2123-01 100X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BG7B2

Quant Time: Apr 24 03:47:20 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM042121WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Apr 24 03:42:40 2021
Response via : Initial Calibration

