

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042324\
 Data File : VV035351.D
 Acq On : 23 Apr 2024 11:18
 Operator : SY/MD
 Sample : P2157-01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :

Quant Time: Apr 23 23:27:54 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Apr 23 23:26:51 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	319298	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	323180	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	148659	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	187555	60.823	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery = 121.640%			
7) Chloroethane-d5	1.532	69	151907	56.717	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery = 113.440%			
11) 1,1-Dichloroethene-d2	2.060	65	77021	56.980	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery = 113.960%			
21) 2-Butanone-d5	3.786	46	201982	110.443	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery = 110.440%			
24) Chloroform-d	4.246	84	321413	61.994	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 123.980%			
26) 1,2-Dichloroethane-d4	4.941	65	199761	60.687	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 121.380%			
32) Benzene-d6	4.957	84	639874	58.943	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 117.880%			
36) 1,2-Dichloropropane-d6	5.986	67	206517	59.908	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery = 119.820%			
41) Toluene-d8	7.240	98	554714	56.739	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 113.480%			
43) trans-1,3-Dichloroprop...	7.551	79	88857	53.022	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery = 106.040%			
47) 2-Hexanone-d5	8.024	63	110491	84.775	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery = 84.780%			
56) 1,1,2,2-Tetrachloroeth...	10.149	84	279017	58.776	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery = 117.560%			
66) 1,2-Dichlorobenzene-d4	11.558	152	220511	66.066	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 132.140%#			
Target Compounds						Qvalue
10) 1,1,2-Trichloro-1,2,2-...	2.060	101	2173	0.867	ug/L #	22
12) 1,1-Dichloroethene	2.060	96	2004	0.836	ug/L #	1
13) Acetone	2.118	43	8662	5.159	ug/L	98
48) 2-Hexanone	8.082	43	29295	9.263	ug/L #	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042324\
Data File : VV035351.D
Acq On : 23 Apr 2024 11:18
Operator : SY/MD
Sample : P2157-01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :

Quant Time: Apr 23 23:27:54 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Apr 23 23:26:51 2024
Response via : Initial Calibration





#10

1,1,2-Trichloro-1,2,2-trifluoroethane

Concen: 0.867 ug/L

RT: 2.060 min Scan# 317

Delta R.T. -0.009 min

Lab File: VV035351.D

Acq: 23 Apr 2024 11:18

Instrument :

MSVOA_V

ClientSampleId :





Tgt Ion:101 Resp: 2173

Ion Ratio Lower Upper

101 100

85 0.0 35.3 52.9#

151 0.0 55.0 82.4#



#12

1,1-Dichloroethene

Concen: 0.836 ug/L

RT: 2.060 min Scan# 317

Delta R.T. -0.009 min

Lab File: VV035351.D

Acq: 23 Apr 2024 11:18




Tgt Ion: 96 Resp: 2004


Ion Ratio Lower Upper

96 100

61 1345.0 131.5 244.3#

63 11815.7 111.4 206.8#





#13

Acetone

Concen: 5.159 ug/L

RT: 2.118 min Scan# 34

Delta R.T. 0.000 min

Lab File: VV035351.D

Acq: 23 Apr 2024 11:18

Instrument :

MSVOA_V





ClientSampleId :

Tgt Ion: 43 Resp: 8662

Ion Ratio Lower Upper

43 100

58 34.1 0.0 66.0



#48

2-Hexanone

Concen: 9.263 ug/L

RT: 8.082 min Scan# 2190

Delta R.T. 0.010 min

Lab File: VV035351.D

Acq: 23 Apr 2024 11:18

Tgt Ion: 43 Resp: 29295

Ion Ratio Lower Upper

43 100

58 56.8 45.4 68.2

57 13.6 16.2 24.4#

100 9.6 10.1 15.1#

