

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW042324\
 Data File : VW035351.D
 Acq On : 23 Apr 2024 11:18
 Operator : SY/MD
 Sample : P2157-01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 CORJ2

Quant Time: Apr 23 23:27:54 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Apr 23 23:26:51 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	319298	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	323180	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	148659	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	187555	60.823	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery = 121.640%			
7) Chloroethane-d5	1.532	69	151907	56.717	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery = 113.440%			
11) 1,1-Dichloroethene-d2	2.060	65	77021	56.980	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery = 113.960%			
21) 2-Butanone-d5	3.786	46	201982	110.443	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery = 110.440%			
24) Chloroform-d	4.246	84	321413	61.994	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 123.980%			
26) 1,2-Dichloroethane-d4	4.941	65	199761	60.687	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 121.380%			
32) Benzene-d6	4.957	84	639874	58.943	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 117.880%			
36) 1,2-Dichloropropane-d6	5.986	67	206517	59.908	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery = 119.820%			
41) Toluene-d8	7.240	98	554714	56.739	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 113.480%			
43) trans-1,3-Dichloroprop...	7.551	79	88857	53.022	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery = 106.040%			
47) 2-Hexanone-d5	8.024	63	110491	84.775	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery = 84.780%			
56) 1,1,2,2-Tetrachloroeth...	10.149	84	279017	58.776	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery = 117.560%			
66) 1,2-Dichlorobenzene-d4	11.558	152	220511	66.066	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 132.140%#			
Target Compounds					Qvalue	
13) Acetone	2.118	43	8662	5.159	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042324\
Data File : VV035351.D
Acq On : 23 Apr 2024 11:18
Operator : SY/MD
Sample : P2157-01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
CORJ2

Quant Time: Apr 23 23:27:54 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Apr 23 23:26:51 2024
Response via : Initial Calibration

