

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042324\
 Data File : VV035371.D
 Acq On : 23 Apr 2024 19:24
 Operator : SY/MD
 Sample : P2158-10DL 50X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 MCORE0DL

Quant Time: Apr 23 23:32:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Apr 23 23:26:51 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	410684	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	412823	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	197391	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	173030	43.626	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	87.260%
7) Chloroethane-d5	1.532	69	142692	41.421	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	82.840%
11) 1,1-Dichloroethene-d2	2.060	65	72567	41.739	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	83.480%
21) 2-Butanone-d5	3.796	46	201134	85.506	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	85.510%
24) Chloroform-d	4.246	84	315181	47.265	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.520%
26) 1,2-Dichloroethane-d4	4.941	65	195364	46.144	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.280%
32) Benzene-d6	4.957	84	622507	44.892	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.780%
36) 1,2-Dichloropropane-d6	5.986	67	201565	45.775	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	91.540%
41) Toluene-d8	7.243	98	533480	42.718	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	85.440%
43) trans-1,3-Dichloroprop...	7.551	79	84624	39.531	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	79.060%
47) 2-Hexanone-d5	8.027	63	107163	64.368	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	64.370%
56) 1,1,2,2-Tetrachloroeth...	10.149	84	279093	46.026	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	92.060%
66) 1,2-Dichlorobenzene-d4	11.558	152	223746	50.486	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.980%
Target Compounds						
					Qvalue	
33) Benzene	5.011	78	227739	16.918	ug/L	100
51) Chlorobenzene	8.812	112	796519	87.202	ug/L	98
64) 1,3-Dichlorobenzene	11.120	146	32457	4.881	ug/L	99
65) 1,4-Dichlorobenzene	11.207	146	277081	40.972	ug/L	99
67) 1,2-Dichlorobenzene	11.577	146	290426	43.840	ug/L	98
70) 1,2,4-trichlorobenzene	13.197	180	108706	24.692	ug/L	99
72) 1,2,3-Trichlorobenzene	13.680	180	30849	7.109	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042324\
Data File : VV035371.D
Acq On : 23 Apr 2024 19:24
Operator : SY/MD
Sample : P2158-10DL 50X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
MC0RE0DL

Quant Time: Apr 23 23:32:12 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Apr 23 23:26:51 2024
Response via : Initial Calibration

