

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW042324\
 Data File : VW035348.D
 Acq On : 23 Apr 2024 10:05
 Operator : SY/MD
 Sample : VW0423WBL01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK273

Quant Time: Apr 23 23:27:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Apr 23 23:26:51 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	422069	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	424753	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	196523	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.282	65	191673	47.023	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	94.040%
7) Chloroethane-d5	1.532	69	155642	43.962	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	87.920%
11) 1,1-Dichloroethene-d2	2.060	65	78875	44.144	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	88.280%
21) 2-Butanone-d5	3.789	46	205622	85.056	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	85.060%
24) Chloroform-d	4.246	84	327742	47.822	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.640%
26) 1,2-Dichloroethane-d4	4.941	65	204873	47.085	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.160%
32) Benzene-d6	4.957	84	658860	46.179	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.360%
36) 1,2-Dichloropropane-d6	5.986	67	210858	46.540	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	93.080%
41) Toluene-d8	7.243	98	561872	43.728	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	87.460%
43) trans-1,3-Dichloroprop...	7.551	79	86232	39.151	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	78.300%
47) 2-Hexanone-d5	8.024	63	114197	66.666	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	66.670%
56) 1,1,2,2-Tetrachloroeth...	10.149	84	287776	46.125	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	92.240%
66) 1,2-Dichlorobenzene-d4	11.558	152	223101	50.563	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.120%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042324\
Data File : VV035348.D
Acq On : 23 Apr 2024 10:05
Operator : SY/MD
Sample : VV0423WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK273

Quant Time: Apr 23 23:27:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Apr 23 23:26:51 2024
Response via : Initial Calibration

