

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV042419\
 Data File : VV010463.D
 Acq On : 24 Apr 2019 15:27
 Operator : SY/MD
 Sample : K2436-10
 Misc : 25mL/MSVOA V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0XA4

Quant Time: Apr 25 03:02:03 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Apr 25 02:19:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	228150	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	230502	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	110748	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	69194	5.20	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	104.00%
7) Chloroethane-d5	1.58	69	54208	5.44	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	108.80%
11) 1,1-Dichloroethene-d2	2.13	63	102976	4.00	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	80.00%
20) 2-Butanone-d5	3.95	46	151826	53.84	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	107.68%
24) Chloroform-d	4.40	84	142380	5.17	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.40%
26) 1,2-Dichloroethane-d4	5.08	65	73548	5.35	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.00%
32) Benzene-d6	5.10	84	273472	4.97	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.40%
36) 1,2-Dichloropropane-d6	6.12	67	79824	4.95	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.00%
41) Toluene-d8	7.36	98	258899	4.81	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.20%
43) trans-1,3-Dichloropropene-	7.67	79	28509	4.48	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.60%
46) 2-Hexanone-d5	8.14	63	127073	52.30	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	104.60%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	60675	5.05	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	101.00%
64) 1,2-Dichlorobenzene-d4	11.67	152	106857	5.22	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	104.40%

Target Compounds

					Ovalue
9) Trichlorofluoromethane	1.77	101	53926	1.871 ug/L	97
13) Acetone	2.19	43	27912	16.451 ug/L	98
22) cis-1,2-Dichloroethene	3.96	96	68380	4.377 ug/L	99
34) Trichloroethene	5.96	95	244509	14.380 ug/L	95
47) Tetrachloroethene	8.02	164	4543	0.299 ug/L	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV042419\
Data File : VV010463.D
Acq On : 24 Apr 2019 15:27
Operator : SY/MD
Sample : K2436-10
Misc : 25mL/MSVOA V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0XA4

Quant Time: Apr 25 03:02:03 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Apr 25 02:19:56 2019
Response via : Initial Calibration

