

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV042419\
 Data File : VV010457.D
 Acq On : 24 Apr 2019 12:43
 Operator : SY/MD
 Sample : VSTDICV005
 Misc : 25mL/MSVOA V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VICV38

Quant Time: Apr 25 02:22:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Apr 25 02:19:56 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Difluorobenzene	1.000	1.000	0.0	114	0.00
2 T	Dichlorodifluoromethane	0.390	0.372	4.6	103	0.00
3 T	Chloromethane	0.315	0.280	11.1	96	0.00
4 S	Vinyl Chloride-d3	0.291	0.248	14.8	95	0.00
5 T	Vinyl chloride	0.341	0.313	8.2	99	0.00
6 T	Bromomethane	0.224	0.209	6.7	99	0.00
7 S	Chloroethane-d5	0.218	0.194	11.0	97	0.00
8 T	Chloroethane	0.191	0.177	7.3	102	0.00
9 T	Trichlorofluoromethane	0.632	0.572	9.5	97	0.00
10 T	1,1,2-Trichloro-1,2,2-trifl	0.332	0.319	3.9	101	0.00
11 S	1,1-Dichloroethene-d2	0.564	0.505	10.5	97	0.00
12 T	1,1-Dichloroethene	0.301	0.277	8.0	99	0.00
13 T	Acetone	0.037	0.034	8.1	97	0.03
14 T	Carbon disulfide	0.814	0.750	7.9	100	0.00
15 T	Methyl Acetate	0.096	0.084	12.5	102	0.01
16 T	Methylene chloride	0.329	0.282	14.3	103	0.00
17 T	Methyl tert-butyl Ether	0.708	0.641	9.5	98	0.00
18 T	trans-1,2-Dichloroethene	0.329	0.308	6.4	102	0.00
19 T	1,1-Dichloroethane	0.524	0.488	6.9	101	0.00
20 S	2-Butanone-d5	0.062	0.056	9.7	100	0.03
21 T	2-Butanone	0.065	0.063	3.1	100	0.03
22 T	cis-1,2-Dichloroethene	0.342	0.315	7.9	101	0.00
23 T	Bromochloromethane	0.149	0.142	4.7	105	0.00
24 S	Chloroform-d	0.604	0.539	10.8	99	0.00
25 T	Chloroform	0.617	0.578	6.3	100	0.00
26 S	1,2-Dichloroethane-d4	0.301	0.271	10.0	100	0.00
27 T	1,2-Dichloroethane	0.362	0.334	7.7	102	0.00
28 I	Chlorobenzene-d5	1.000	1.000	0.0	111	0.00
29 T	1,1,1-Trichloroethane	0.582	0.562	3.4	102	0.00
30 T	Cyclohexane	0.491	0.474	3.5	102	0.00
31 T	Carbon tetrachloride	0.529	0.514	2.8	98	0.00
32 S	Benzene-d6	1.193	1.085	9.1	100	0.00
33 T	Benzene	1.278	1.222	4.4	100	0.00
34 T	Trichloroethene	0.369	0.353	4.3	101	0.00
35 T	Methylcyclohexane	0.554	0.529	4.5	97	0.00
36 S	1,2-Dichloropropane-d6	0.350	0.318	9.1	98	0.00
37 T	1,2-Dichloropropane	0.310	0.295	4.8	104	0.00
38 T	Bromodichloromethane	0.418	0.400	4.3	101	0.00
39 T	cis-1,3-Dichloropropene	0.460	0.445	3.3	101	0.00
40 T	4-Methyl-2-pentanone	0.175	0.168	4.0	101	0.00
41 S	Toluene-d8	1.168	1.090	6.7	100	0.00
42 T	Toluene	1.455	1.382	5.0	100	0.00
43 S	trans-1,3-Dichloropropene-d	0.138	0.124	10.1	100	0.00
44 T	trans-1,3-Dichloropropene	0.355	0.340	4.2	100	0.00
45 T	1,1,2-Trichloroethane	0.227	0.216	4.8	100	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV042419\
 Data File : VV010457.D
 Acq On : 24 Apr 2019 12:43
 Operator : SY/MD
 Sample : VSTDICV005
 Misc : 25mL/MSVOA V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VICV38

Quant Time: Apr 25 02:22:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Apr 25 02:19:56 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 S	2-Hexanone-d5	0.053	0.048	9.4	97	0.00
47 T	Tetrachloroethene	0.329	0.303	7.9	99	0.00
48 T	2-Hexanone	0.122	0.123	-0.8	103	0.00
49 T	Dibromochloromethane	0.294	0.276	6.1	99	0.00
50 T	1,2-Dibromoethane	0.220	0.207	5.9	98	0.00
51 T	Chlorobenzene	0.988	0.927	6.2	102	0.00
52 T	Ethylbenzene	1.616	1.550	4.1	101	0.00
53 T	m,p-xylene	0.636	0.609	4.2	98	0.00
54 T	o-xylene	0.624	0.585	6.3	99	0.00
55 T	Styrene	1.011	0.986	2.5	100	0.00
56 T	Isopropylbenzene	1.661	1.591	4.2	98	0.00
57 S	1,1,2,2-Tetrachloroethane-d	0.260	0.239	8.1	100	0.00
58 T	1,1,2,2-Tetrachloroethane	0.261	0.244	6.5	98	0.00
59	1,2,3-Trichloropropane	0.192	0.177	7.8	96	0.00
60 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	111	0.00
61 T	Bromoform	0.296	0.276	6.8	102	0.00
62 T	1,3-Dichlorobenzene	1.558	1.493	4.2	102	0.00
63 T	1,4-Dichlorobenzene	1.569	1.496	4.7	103	0.00
64 S	1,2-Dichlorobenzene-d4	0.924	0.819	11.4	102	0.00
65 T	1,2-Dichlorobenzene	1.478	1.391	5.9	101	0.00
66 T	1,2-Dibromo-3-chloropropane	0.080	0.072	10.0	107	0.00
67	1,3,5-Trichlorobenzene	1.276	1.233	3.4	104	0.00
68 T	1,2,4-trichlorobenzene	0.965	0.978	-1.3	110	0.00
69	Naphthalene	1.432	1.430	0.1	113	0.00
70 T	1,2,3-Trichlorobenzene	0.915	0.900	1.6	108	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0