

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV042419\
 Data File : VV010485.D
 Acq On : 25 Apr 2019 01:42
 Operator : SY/MD
 Sample : K2477-20
 Misc : 25mL/MSVOA V/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0XC7

Quant Time: Apr 25 04:42:15 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Apr 25 04:02:34 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	229908	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	234876	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	105214	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	60703	4.53	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	90.60%
7) Chloroethane-d5	1.58	69	49537	4.94	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	98.80%
11) 1,1-Dichloroethene-d2	2.13	63	91945	3.55	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	71.00%
20) 2-Butanone-d5	3.95	46	147238	51.81	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	103.62%
24) Chloroform-d	4.40	84	133138	4.80	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.00%
26) 1,2-Dichloroethane-d4	5.08	65	70890	5.12	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.40%
32) Benzene-d6	5.10	84	275932	4.92	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.40%
36) 1,2-Dichloropropane-d6	6.12	67	81544	4.96	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.20%
41) Toluene-d8	7.36	98	245261	4.47	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.40%
43) trans-1,3-Dichloropropene-	7.67	79	25660	3.95	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	79.00%
46) 2-Hexanone-d5	8.14	63	110859	44.78	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	89.56%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	59512	4.87	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.40%
64) 1,2-Dichlorobenzene-d4	11.67	152	99224	5.10	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.00%

Target Compounds


					Ovalue
9) Trichlorofluoromethane	1.77	101	8734	0.301 ug/L	92
13) Acetone	2.19	43	3718	2.175 ug/L	90
34) Trichloroethene	5.96	95	12448	0.718 ug/L	90
47) Tetrachloroethene	8.02	164	29098	1.881 ug/L	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV042419\
Data File : VV010485.D
Acq On : 25 Apr 2019 01:42
Operator : SY/MD
Sample : K2477-20
Misc : 25mL/MSVOA V/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0XC7

Quant Time: Apr 25 04:42:15 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Apr 25 04:02:34 2019
Response via : Initial Calibration









#9

Trichlorofluoromethane
 Concen: 0.301 ug/L
 RT: 1.77 min Scan# 212
 Delta R.T. 0.00 min
 Lab File: VV010485.D
 Acq: 25 Apr 2019 01:42

Instrument :
 MSVOA_V
 ClientSampleId :
 C0XC7




Tot Ion:101 Resp: 8734
 Ion Ratio Lower Upper
 101 100
 103 72.2 52.6 78.8




#13

Acetone
 Concen: 2.175 ug/L
 RT: 2.19 min Scan# 342
 Delta R.T. -0.00 min
 Lab File: VV010485.D
 Acq: 25 Apr 2019 01:42

Tgt Ion: 43 Resp: 3718
 Ion Ratio Lower Upper
 43 100
 58 35.1 0.0 59.6





#34

Trichloroethene

Concen: 0.718 ug/L

RT: 5.96 min Scan# 1514

Delta R.T. 0.00 min

Lab File: VV010485.D

Acq: 25 Apr 2019 01:42

Instrument :

MSVOA_V

ClientSampled :

C0XC7

Tot Ion: 95 Resp: 12448


Ion Ratio Lower Upper

95 100

97 52.0 45.6 84.6

132 89.3 67.8 126.0

130 106.4 69.0 128.2






Abundance Ion 95.00 (94.70 to 95.70): VV0

Ion 96.95 (96.65 to 97.65): VV0

Ion 131.95 (131.65 to 132.65): V

Ion 129.95 (129.65 to 130.65): V



#47

Tetrachloroethene

Concen: 1.881 ug/L

RT: 8.02 min Scan# 2155

Delta R.T. 0.00 min

Lab File: VV010485.D

Acq: 25 Apr 2019 01:42

Tgt Ion:164 Resp: 29098


Ion Ratio Lower Upper

164 100

129 96.3 69.7 129.4

131 96.3 63.8 118.6

166 124.1 87.4 162.2



Abundance Ion 163.90 (163.60 to 164.60): V

Ion 128.95 (128.65 to 129.65): V

Ion 130.95 (130.65 to 131.65): V

Ion 165.90 (165.60 to 166.60): V

