

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV042420\
 Data File : VV015674.D
 Acq On : 24 Apr 2020 15:22
 Operator : SY/MD
 Sample : L2416-06
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 ETKP5

Quant Time: Apr 25 02:36:00 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Apr 25 00:15:39 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	125251	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	119675	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	58875	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	26654	4.07	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	81.40%
7) Chloroethane-d5	1.58	69	29342	4.65	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	93.00%
11) 1,1-Dichloroethene-d2	2.12	63	45197	3.15	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	63.00%
20) 2-Butanone-d5	3.91	46	103162	49.08	ug/L	-0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	98.16%
24) Chloroform-d	4.38	84	70820	4.64	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.80%
26) 1,2-Dichloroethane-d4	5.06	65	41954	4.92	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.40%
32) Benzene-d6	5.08	84	133824	4.68	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.60%
36) 1,2-Dichloropropane-d6	6.09	67	42942	4.70	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.00%
41) Toluene-d8	7.34	98	123198	4.54	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.80%
43) trans-1,3-Dichloropropene-	7.64	79	14482	4.10	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	82.00%
46) 2-Hexanone-d5	8.11	63	77038	44.76	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	89.52%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	30113	4.43	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	88.60%
64) 1,2-Dichlorobenzene-d4	11.65	152	48201	4.82	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.40%

Target Compounds

					Ovalue
13) Acetone	2.18	43	13579	9.305	ug/L 99
16) Methylene chloride	2.52	84	1441	0.160	ug/L 94
27) 1,2-Dichloroethane	5.16	62	4283	0.383	ug/L # 95
52) Ethylbenzene	9.04	91	2957	0.070	ug/L 93
53) m,p-xylene	9.16	106	3741	0.242	ug/L 87
54) o-xylene	9.57	106	1845	0.123	ug/L 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV042420\
Data File : VV015674.D
Acq On : 24 Apr 2020 15:22
Operator : SY/MD
Sample : L2416-06
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
ETKP5

Quant Time: Apr 25 02:36:00 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Apr 25 00:15:39 2020
Response via : Initial Calibration

