

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW042423\
 Data File : VW030801.D
 Acq On : 24 Apr 2023 18:34
 Operator : SY/MD
 Sample : 02479-17
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EW972

Quant Time: Apr 25 03:23:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM041723WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Apr 22 00:43:23 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	234324	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.792	117	226428	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	110135	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.282	65	52488	30.556	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	61.120%
7) Chloroethane-d5	1.539	69	49780	34.582	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	69.160%#
11) 1,1-Dichloroethene-d2	2.066	65	27654	33.947	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	67.900%
21) 2-Butanone-d5	3.806	46	87724	92.977	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	92.980%
24) Chloroform-d	4.262	84	121757	39.112	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	78.220%
26) 1,2-Dichloroethane-d4	4.950	65	85467	42.703	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.400%
32) Benzene-d6	4.969	84	210798	39.376	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	78.760%
36) 1,2-Dichloropropane-d6	5.998	67	66760	40.147	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	80.300%
41) Toluene-d8	7.252	98	183722	36.552	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	73.100%#
43) trans-1,3-Dichloroprop...	7.561	79	31477	37.556	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	75.120%
47) 2-Hexanone-d5	8.034	63	55730	85.073	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	85.070%
56) 1,1,2,2-Tetrachloroeth...	10.159	84	103024	42.696	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	85.400%
66) 1,2-Dichlorobenzene-d4	11.570	152	78778	45.002	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042423\
Data File : VV030801.D
Acq On : 24 Apr 2023 18:34
Operator : SY/MD
Sample : 02479-17
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EW972

Quant Time: Apr 25 03:23:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM041723WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Apr 22 00:43:23 2023
Response via : Initial Calibration

