

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW042423\
 Data File : VW030799.D
 Acq On : 24 Apr 2023 17:46
 Operator : SY/MD
 Sample : 02419-19
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EW932

Quant Time: Apr 25 03:23:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM041723WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Apr 22 00:43:23 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	231178	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.792	117	222183	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	108239	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	51630	30.465	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	60.940%
7) Chloroethane-d5	1.539	69	49981	35.194	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	70.380%
11) 1,1-Dichloroethene-d2	2.066	65	27988	34.824	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	69.640%
21) 2-Butanone-d5	3.809	46	87992	94.530	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	94.530%
24) Chloroform-d	4.262	84	122831	39.994	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	79.980%
26) 1,2-Dichloroethane-d4	4.953	65	86494	43.805	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.600%
32) Benzene-d6	4.969	84	216590	41.231	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	82.460%
36) 1,2-Dichloropropane-d6	5.998	67	68488	41.973	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	83.940%
41) Toluene-d8	7.252	98	188194	38.157	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	76.320%#
43) trans-1,3-Dichloroprop...	7.561	79	33598	40.852	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	81.700%
47) 2-Hexanone-d5	8.034	63	55822	86.842	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	86.840%
56) 1,1,2,2-Tetrachloroeth...	10.159	84	101781	42.987	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	85.980%
66) 1,2-Dichlorobenzene-d4	11.567	152	82316	47.846	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.700%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042423\
Data File : VV030799.D
Acq On : 24 Apr 2023 17:46
Operator : SY/MD
Sample : 02419-19
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EW932

Quant Time: Apr 25 03:23:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM041723WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Apr 22 00:43:23 2023
Response via : Initial Calibration

