


Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042519\
Data File : VV010491.D
Acq On : 25 Apr 2019 09:57
Operator : SY/MD
Sample : BFB37
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFB37

Integration File: rteint.p

Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042419WMA.M
Title : TRACE VOA SOM01.0
Last Update : Fri Apr 26 05:08:05 2019



AutoFind: Scans 2808, 2809, 2810; Background Corrected with Scan 2795

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.5	10943	PASS
75	95	30	80	47.5	29728	PASS
95	95	100	100	100.0	62592	PASS
96	95	5	9	7.3	4575	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	91.4	57213	PASS
175	174	5	9	7.6	4363	PASS
176	174	95	101	97.9	55992	PASS
177	176	5	9	6.4	3608	PASS

Scan 2795 (10.077 min): VV010491.D
BFB37

m/z	Abundance
39.90	1689.0
43.90	289.0

Instrument :
MSVOA_V
ClientSampleId :
BFB37

m/z	Abundance
36.00	628.0
37.00	2942.0
37.90	2527.0
39.00	1295.0
39.90	1966.0
43.90	743.0
45.00	558.0
45.90	117.0
46.90	496.0
48.00	439.0
49.00	2279.0
50.00	11075.0
51.00	3180.0
52.10	122.0
53.10	112.0
55.90	1202.0
57.00	1731.0
59.90	752.0
61.00	3455.0
62.00	3399.0
63.10	2680.0
64.10	515.0
64.90	802.0
66.90	112.0
68.00	6599.0
69.00	7399.0
69.90	633.0
71.90	505.0
73.10	2663.0
74.00	11272.0
75.00	30368.0
76.00	3211.0
77.10	523.0
78.10	230.0
78.90	2912.0
79.90	713.0
80.90	2810.0
81.80	579.0
85.90	136.0
86.90	2717.0
88.10	2150.0
90.90	342.0
92.00	2192.0
93.00	3260.0
94.00	8122.0
95.00	63872.0
96.00	4736.0
103.70	336.0
104.80	217.0
106.00	343.0
112.70	103.0
115.80	285.0
116.90	558.0
118.00	364.0
118.80	494.0
127.90	257.0
128.80	117.0
129.90	317.0
131.10	119.0
135.00	141.0
137.10	232.0
141.00	826.0
142.90	1397.0
145.00	312.0
148.00	224.0
154.90	150.0
157.00	217.0
160.80	148.0
171.00	172.0
172.00	1355.0
174.00	57944.0
175.00	4415.0
176.00	55080.0
177.00	3601.0

Instrument :
MSVOA_V
ClientSampleId :
BFB37

m/z	Abundance
36.00	624.0
37.00	2930.0
38.00	2421.0
39.00	1340.0
39.90	2089.0
41.10	114.0
43.90	782.0
45.00	651.0
46.90	583.0
48.00	453.0
49.00	2443.0
50.00	11429.0
51.00	3387.0
52.10	110.0
53.00	104.0
55.90	1031.0
57.10	1931.0
59.90	689.0
61.00	3207.0
62.00	3230.0
63.10	2393.0
64.10	446.0
64.90	873.0
67.10	162.0
68.00	6955.0
69.00	7388.0
70.00	649.0
71.90	520.0
73.00	2802.0
74.00	11350.0
75.00	30272.0
76.00	3277.0
77.00	468.0
78.10	204.0
78.90	2888.0
79.90	752.0
81.00	2985.0
81.80	629.0
86.90	2538.0
88.10	2370.0
90.80	304.0
92.00	2244.0
93.00	3518.0
94.00	8599.0
95.00	64440.0
96.00	4753.0
103.80	308.0
104.80	207.0
106.00	415.0
110.70	118.0
112.80	117.0
115.90	320.0
116.90	579.0
118.00	400.0
118.80	566.0
128.00	250.0
130.00	344.0
131.10	136.0
135.00	148.0
137.00	219.0
141.00	779.0
142.90	1420.0
144.90	330.0
145.90	195.0
148.00	246.0
150.10	109.0
155.00	152.0
157.00	222.0
158.70	111.0
160.80	200.0
171.00	182.0
172.00	1270.0
174.00	58808.0
175.00	4572.0
176.00	57944.0
177.00	3750.0

Instrument :
MSVOA_V
ClientSampleId :
BFB37

m/z	Abundance
35.90	505.0
37.00	2671.0
38.00	2454.0
39.00	1158.0
39.90	2121.0
41.00	121.0
43.90	704.0
45.00	622.0
47.00	678.0
48.00	377.0
48.90	2434.0
50.00	10325.0
51.00	3240.0
52.00	128.0
55.10	102.0
56.00	917.0
57.10	2068.0
59.90	528.0
61.00	2759.0
62.00	3009.0
63.10	2000.0
64.00	382.0
64.90	787.0
67.00	179.0
68.00	6539.0
69.00	6607.0
70.00	660.0
71.90	404.0
73.00	2725.0
74.00	10577.0
75.00	28544.0
76.00	2900.0
77.00	345.0
78.10	217.0
78.90	2600.0
79.90	716.0
81.00	2807.0
81.80	618.0
87.00	2153.0
87.90	2238.0
90.80	298.0
92.00	2018.0
93.00	3566.0
94.00	8266.0
95.00	59464.0
96.00	4236.0
103.90	301.0
104.80	124.0
105.90	416.0
110.70	124.0
112.80	106.0
115.90	288.0
117.00	528.0
118.10	333.0
118.90	488.0
128.00	278.0
130.00	343.0
131.00	123.0
135.00	126.0
136.90	182.0
141.00	707.0
142.00	114.0
142.90	1300.0
143.90	106.0
144.90	317.0
145.90	224.0
147.90	271.0
155.00	172.0
157.00	183.0
160.80	166.0
171.00	167.0
172.00	1037.0
174.00	54888.0
175.00	4103.0
176.00	54952.0
177.00	3475.0

Instrument :
MSVOA_V
ClientSampleId :
BFB37