

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV042520\
 Data File : VV015714.D
 Acq On : 25 Apr 2020 23:18
 Operator : SY/MD
 Sample : L2413-21MSD
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 ETKM8MSD

Quant Time: Apr 27 05:59:22 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Apr 27 01:12:28 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	128979	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	126040	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	59570	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	36397	5.40	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	108.00%
7) Chloroethane-d5	1.58	69	27064	4.17	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	83.40%
11) 1,1-Dichloroethene-d2	2.13	63	82493	5.59	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	111.80%
20) 2-Butanone-d5	3.92	46	109753	50.71	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	101.42%
24) Chloroform-d	4.38	84	71141	4.53	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.60%
26) 1,2-Dichloroethane-d4	5.06	65	40638	4.62	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.40%
32) Benzene-d6	5.08	84	132171	4.39	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.80%
36) 1,2-Dichloropropane-d6	6.10	67	43717	4.55	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	91.00%
41) Toluene-d8	7.34	98	120300	4.21	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.20%
43) trans-1,3-Dichloropropene-	7.65	79	11419	3.07	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	61.40%
46) 2-Hexanone-d5	8.11	63	82636	45.59	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	91.18%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	31508	4.40	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	88.00%
64) 1,2-Dichlorobenzene-d4	11.65	152	46601	4.60	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	92.00%

Target Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
5) Vinyl chloride	1.32	62	1466694	143.687	ug/L	100
12) 1,1-Dichloroethene	2.13	96	67459	8.986	ug/L #	85
18) trans-1,2-Dichloroethene	2.78	96	96897	11.528	ug/L	99
22) cis-1,2-Dichloroethene	3.94	96	11541542	1241.901	ug/L	95
30) Cyclohexane	4.71	56	9788	0.641	ug/L	95
33) Benzene	5.13	78	197731	5.491	ug/L	100
34) Trichloroethene	5.94	95	2060088	207.153	ug/L	98
35) Methylcyclohexane	6.15	83	4332	0.284	ug/L	88
42) Toluene	7.41	91	210195	5.463	ug/L	99
47) Tetrachloroethene	8.00	164	3666115	511.993	ug/L	97
51) Chlorobenzene	8.90	112	132683	5.264	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV042520\
Data File : VV015714.D
Acq On : 25 Apr 2020 23:18
Operator : SY/MD
Sample : L2413-21MSD
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
ETKM8MSD

Quant Time: Apr 27 05:59:22 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Apr 27 01:12:28 2020
Response via : Initial Calibration

