

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV042520\
 Data File : VV015693.D
 Acq On : 25 Apr 2020 14:46
 Operator : SY/MD
 Sample : L2388-20
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFSD3

Quant Time: Apr 27 02:31:47 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Apr 27 01:12:28 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	121574	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	117650	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	55737	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	26478	4.17	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	83.40%
7) Chloroethane-d5	1.58	69	28793	4.70	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	94.00%
11) 1,1-Dichloroethene-d2	2.12	63	44189	3.18	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	63.60%
20) 2-Butanone-d5	3.91	46	102318	50.16	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	100.32%
24) Chloroform-d	4.38	84	71710	4.84	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.80%
26) 1,2-Dichloroethane-d4	5.06	65	42763	5.16	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.20%
32) Benzene-d6	5.07	84	134878	4.80	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.00%
36) 1,2-Dichloropropane-d6	6.09	67	44213	4.93	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.60%
41) Toluene-d8	7.34	98	120347	4.51	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.20%
43) trans-1,3-Dichloropropene-	7.64	79	12524	3.61	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	72.20%
46) 2-Hexanone-d5	8.11	63	74257	43.89	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	87.78%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	31206	4.67	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.40%
64) 1,2-Dichlorobenzene-d4	11.65	152	47512	5.01	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.20%


Target Compounds					Ovalue
13) Acetone	2.19	43	2490	1.758 ug/L	86


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV042520\
Data File : VV015693.D
Acq On : 25 Apr 2020 14:46
Operator : SY/MD
Sample : L2388-20
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFSD3

Quant Time: Apr 27 02:31:47 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Apr 27 01:12:28 2020
Response via : Initial Calibration





#13

Acetone

Concen: 1.758 ug/L

RT: 2.19 min Scan# 342

Delta R.T. -0.02 min

Lab File: VV015693.D

Acq: 25 Apr 2020 14:46

Instrument :
MSVOA_V
ClientSampleId :
BFSD3

Tot Ion: 43 Resp: 2490

Ion Ratio Lower Upper

43 100

58 21.3 0.0 57.6

