

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042523\
 Data File : VV030858.D
 Acq On : 26 Apr 2023 00:19
 Operator : SY/MD
 Sample : 02474-09
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AT1

Quant Time: Apr 26 02:58:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM041723WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Apr 22 00:43:23 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	199993	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.792	117	196711	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	98525	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	63738	43.474	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	86.940%	
7) Chloroethane-d5	1.535	69	54962	44.736	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	89.480%	
11) 1,1-Dichloroethene-d2	2.066	65	33389	48.023	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	96.040%	
21) 2-Butanone-d5	3.796	46	86730	107.703	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	107.700%	
24) Chloroform-d	4.259	84	139077	52.345	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	104.700%	
26) 1,2-Dichloroethane-d4	4.950	65	98056	57.404	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	114.800%	
32) Benzene-d6	4.969	84	251956	54.175	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	108.340%	
36) 1,2-Dichloropropane-d6	5.998	67	77062	53.343	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	106.680%	
41) Toluene-d8	7.252	98	227185	52.027	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	104.060%	
43) trans-1,3-Dichloroprop...	7.561	79	37013	50.832	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	101.660%	
47) 2-Hexanone-d5	8.034	63	51261	90.072	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	90.070%	
56) 1,1,2,2-Tetrachloroeth...	10.162	84	105618	50.384	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	100.760%	
66) 1,2-Dichlorobenzene-d4	11.570	152	93548	59.736	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	119.480%	
Target Compounds						
13) Acetone	2.121	43	4616	3.704	ug/L	88
30) 1,1,1-Trichloroethane	4.526	97	5435	1.966	ug/L #	90
34) Trichloroethene	5.844	95	19810	12.750	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042523\
Data File : VV030858.D
Acq On : 26 Apr 2023 00:19
Operator : SY/MD
Sample : 02474-09
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
COAT1

Quant Time: Apr 26 02:58:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM041723WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Apr 22 00:43:23 2023
Response via : Initial Calibration

