

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW042524\
 Data File : VW035413.D
 Acq On : 25 Apr 2024 18:53
 Operator : SY/MD
 Sample : P2203-07
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK282

Quant Time: Apr 26 01:40:04 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Apr 26 01:37:05 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	379434	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	380708	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	159301	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	192174	52.444	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	104.880%	
7) Chloroethane-d5	1.529	69	143020	44.936	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	89.880%	
11) 1,1-Dichloroethene-d2	2.053	65	79376	49.416	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	98.840%	
21) 2-Butanone-d5	3.796	46	189016	86.973	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	86.970%	
24) Chloroform-d	4.246	84	315944	51.281	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	102.560%	
26) 1,2-Dichloroethane-d4	4.937	65	195225	49.909	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	99.820%	
32) Benzene-d6	4.957	84	641405	50.156	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	100.320%	
36) 1,2-Dichloropropane-d6	5.985	67	201262	49.561	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	99.120%	
41) Toluene-d8	7.239	98	541925	47.055	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	94.120%	
43) trans-1,3-Dichloroprop...	7.551	79	84617	42.862	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	85.720%	
47) 2-Hexanone-d5	8.027	63	99296	64.673	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	64.670%	
56) 1,1,2,2-Tetrachloroeth...	10.149	84	268407	47.997	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	96.000%	
66) 1,2-Dichlorobenzene-d4	11.558	152	193788	54.181	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	108.360%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042524\
Data File : VV035413.D
Acq On : 25 Apr 2024 18:53
Operator : SY/MD
Sample : P2203-07
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK282

Quant Time: Apr 26 01:40:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Apr 26 01:37:05 2024
Response via : Initial Calibration

