

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042524\
 Data File : VV035402.D
 Acq On : 25 Apr 2024 13:44
 Operator : SY/MD
 Sample : VV0425WBL01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK275

Quant Time: Apr 26 01:37:28 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Apr 26 01:37:05 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	386642	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	380246	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	165107	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	173706	46.520	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	93.040%	
7) Chloroethane-d5	1.519	69	78585	24.230	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	48.460%#	
11) 1,1-Dichloroethene-d2	2.053	65	74166	45.311	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	90.620%	
21) 2-Butanone-d5	3.789	46	183884	83.034	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	83.030%	
24) Chloroform-d	4.243	84	317402	50.557	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	101.120%	
26) 1,2-Dichloroethane-d4	4.934	65	196747	49.361	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	98.720%	
32) Benzene-d6	4.953	84	645631	50.548	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	101.100%	
36) 1,2-Dichloropropane-d6	5.982	67	202516	49.931	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	99.860%	
41) Toluene-d8	7.239	98	551004	47.901	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	95.800%	
43) trans-1,3-Dichloroprop...	7.551	79	84951	43.083	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	86.160%	
47) 2-Hexanone-d5	8.027	63	99298	64.753	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	64.750%	
56) 1,1,2,2-Tetrachloroeth...	10.149	84	271939	48.688	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	97.380%	
66) 1,2-Dichlorobenzene-d4	11.558	152	203710	54.953	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	109.900%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042524\
Data File : VV035402.D
Acq On : 25 Apr 2024 13:44
Operator : SY/MD
Sample : VV0425WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK275

Quant Time: Apr 26 01:37:28 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Apr 26 01:37:05 2024
Response via : Initial Calibration

